

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL081522\
 Data File : PL077135.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Aug 2022 22:49
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 08/16/2022
 Supervised By :Ankita Jodhani 08/16/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 16 07:01:11 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL080822.M
 Quant Title : GC Extractables
 QLast Update : Tue Aug 09 09:13:15 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.418	2.635	52331911	136.1E6	53.370	47.546
28)	SA Decachlor...	8.907	7.738	41020593	112.7E6	54.336	50.079
Target Compounds							
2)	A alpha-BHC	3.883	3.128	75053584	203.1E6	54.219	47.891
3)	MA gamma-BHC...	4.218	3.452	69755084	184.7E6	54.550	48.407
4)	MA Heptachlor	4.794	3.785	66448145	172.9E6	54.084	48.184
5)	MB Aldrin	5.134	4.060	62285411	171.2E6	54.895	48.331
6)	B beta-BHC	4.429	3.751	29711938	74157274	52.933	47.462
7)	B delta-BHC	4.673	3.975	63361186	185.5E6	56.908	48.536
8)	B Heptachlo...	5.567	4.559	55477120	146.9E6	54.941	48.336
9)	A Endosulfan I	5.953	4.925	50447574	132.0E6	55.077	48.286
10)	B gamma-Chl...	5.822	4.808	55063931	147.8E6	54.938	48.509
11)	B alpha-Chl...	5.899	4.871	54874181	144.5E6	54.980	48.512
12)	B 4,4'-DDE	6.078	5.065	48513949	141.4E6	55.600	49.024
13)	MA Dieldrin	6.230	5.186	53747448	148.4E6	55.352	49.473m
14)	MA Endrin	6.462	5.461	47125516	128.6E6	57.905	50.618
15)	B Endosulfa...	6.689	5.756	46935669	122.7E6	54.719	46.844
16)	A 4,4'-DDD	6.603	5.617	41169797	112.2E6	54.040	49.796
17)	MA 4,4'-DDT	6.913	5.866	42534135	111.8E6	55.351	49.336
18)	B Endrin al...	6.820	5.935	37622596	89965968	52.180	45.759
19)	B Endosulfa...	7.054	6.158	45105791	112.5E6	54.556	48.762
20)	A Methoxychlor	7.395	6.444	24660043	59497722	55.555	49.465
21)	B Endrin ke...	7.542	6.658	49833827	122.9E6	55.011	48.333
22)	Mirex	8.003	6.835	38341287	92858134	54.571	48.535

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL081522\
Data File : PL077135.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Aug 2022 22:49
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :

ECD_L

ClientSampleId :

PSTDCCC050

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 08/16/2022
Supervised By :Ankita Jodhani 08/16/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 16 07:01:11 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL080822.M
Quant Title : GC Extractables
QLast Update : Tue Aug 09 09:13:15 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

