

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL081625\
 Data File : PL096831.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Aug 2025 18:59
 Operator : AR\AJ
 Sample : Q2814-01
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 TW-84SB-E

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 18 05:56:04 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL072825.M
 Quant Title : GC Extractables
 QLast Update : Fri Aug 08 15:43:38 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.531	2.826	87327776	98184634	27.455	20.550 #
28)	SA Decachlor...	9.004	7.987	39557418	74905790	16.588	17.266
Target Compounds							
9)	A Endosulfan I	0.000	5.177	0	10592134	N.D.	1.915 #
18)	B Endrin al...	0.000	6.169	0	15330795	N.D.	4.217 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL081625\
Data File : PL096831.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Aug 2025 18:59
Operator : AR\AJ
Sample : Q2814-01
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
TW-84SB-E

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 18 05:56:04 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL072825.M
Quant Title : GC Extractables
QLast Update : Fri Aug 08 15:43:38 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.531 min
Delta R.T.: -0.004 min
Response: 87327776
Conc: 27.45 ng/ml

Instrument :
ECD_L
ClientSampleId :
TW-84SB-E




#1 Tetrachloro-m-xylene

R.T.: 2.826 min
Delta R.T.: -0.002 min
Response: 98184634
Conc: 20.55 ng/ml




#9 Endosulfan I

R.T.: 0.000 min
Exp R.T. : 6.044 min
Response: 0
Conc: N.D.



#9 Endosulfan I


R.T.: 5.177 min
Delta R.T.: 0.008 min
Response: 10592134
Conc: 1.91 ng/ml



#18 Endrin aldehyde


R.T.: 0.000 min
Exp R.T.: 6.884 min
Response: 0
Conc: N.D.

Instrument :
ECD_L
ClientSampleId :
TW-84SB-E




#18 Endrin aldehyde

R.T.: 6.169 min
Delta R.T.: -0.010 min
Response: 15330795
Conc: 4.22 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.004 min
Delta R.T.: -0.012 min
Response: 39557418
Conc: 16.59 ng/ml



#28 Decachlorobiphenyl

R.T.: 7.987 min
Delta R.T.: -0.005 min
Response: 74905790
Conc: 17.27 ng/ml