

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL081718\
 Data File : PL039290.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 Aug 2018 23:31
 Operator : AJ\SJ
 Sample : J4500-05
 Misc :
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 18 00:13:35 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL081318.M
 Quant Title : GC Extractables
 QLast Update : Wed Aug 15 15:04:22 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.305	3.907	91044520	39640062	13.024	14.786
28)	SA Decachlor...	7.962	8.896	93743517	38025234	13.347	16.386
Target Compounds							
6)	B beta-BHC	4.217f	0.000	15709699	0	3.382	N.D. #
8)	B Heptachlo...	4.923f	5.774	1545682	7765054	0.162	2.309 #
10)	B gamma-Chl...	5.135f	0.000	9429263	0	0.955	N.D. #
25)	Chlordane-3	5.135	0.000	9429263	0	8.633	N.D. #


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL081718\
Data File : PL039290.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Aug 2018 23:31
Operator : AJ\SJ
Sample : J4500-05
Misc :
ALS Vial : 34 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 18 00:13:35 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL081318.M
Quant Title : GC Extractables
QLast Update : Wed Aug 15 15:04:22 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.305 min
Delta R.T.: -0.001 min
Response: 91044520
Conc: 13.02 ng/ml

Instrument :
ECD_L
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.907 min
Delta R.T.: 0.000 min
Response: 39640062
Conc: 14.79 ng/ml




#6 beta-BHC

R.T.: 4.217 min
Delta R.T.: -0.047 min
Response: 15709699
Conc: 3.38 ng/ml



#6 beta-BHC


R.T.: 0.000 min
Exp R.T. : 4.726 min
Response: 0
Conc: N.D.



#8 Heptachlor epoxide


R.T.: 4.923 min
Delta R.T.: -0.043 min
Response: 1545682
Conc: 0.16 ng/ml

Instrument :
ECD_L
ClientSampleId :




#8 Heptachlor epoxide

R.T.: 5.774 min
Delta R.T.: 0.023 min
Response: 7765054
Conc: 2.31 ng/ml




#10 gamma-Chlordane

R.T.: 5.135 min
Delta R.T.: -0.052 min
Response: 9429263
Conc: 0.95 ng/ml



#10 gamma-Chlordane


R.T.: 0.000 min
Exp R.T.: 5.983 min
Response: 0
Conc: N.D.



#25 Chlordane-3


R.T.: 5.135 min
Delta R.T.: -0.053 min
Response: 9429263
Conc: 8.63 ng/ml

Instrument :
ECD_L
ClientSampleId :




#25 Chlordane-3

R.T.: 0.000 min
Exp R.T. : 5.983 min
Response: 0
Conc: N.D.



#28 Decachlorobiphenyl

R.T.: 7.962 min
Delta R.T.: -0.003 min
Response: 93743517
Conc: 13.35 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.896 min
Delta R.T.: 0.000 min
Response: 38025234
Conc: 16.39 ng/ml