

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL081718\
 Data File : PL039292.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Aug 2018 00:00
 Operator : AJ\SJ
 Sample : J4500-13
 Misc :
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 18 03:12:17 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL081318.M
 Quant Title : GC Extractables
 QLast Update : Wed Aug 15 15:04:22 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.305	3.907	80508845	34366853	11.517	12.819
28)	SA Decachlor...	7.963	8.895	76655230	30902534	10.914	13.317
Target Compounds							
6)	B beta-BHC	4.216f	0.000	15041939	0	3.238	N.D. #
8)	B Heptachlo...	0.000	5.774	0	7382127	N.D.	2.195 #
10)	B gamma-Chl...	5.205	0.000	5803651	0	0.588	N.D. #
11)	B alpha-Chl...	5.205f	0.000	5803651	0	0.604	N.D. #
17)	MA 4,4'-DDT	6.151	0.000	5044372	0	0.665	N.D. #
25)	Chlordane-3	5.205	0.000	5803651	0	5.313	N.D. #
26)	Chlordane-4	5.205	0.000	5803651	0	5.807	N.D. #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL081718\
Data File : PL039292.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Aug 2018 00:00
Operator : AJ\SJ
Sample : J4500-13
Misc :
ALS Vial : 36 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 18 03:12:17 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL081318.M
Quant Title : GC Extractables
QLast Update : Wed Aug 15 15:04:22 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.305 min
Delta R.T.: -0.001 min
Response: 80508845
Conc: 11.52 ng/ml

Instrument :
ECD_L
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.907 min
Delta R.T.: 0.000 min
Response: 34366853
Conc: 12.82 ng/ml




#6 beta-BHC

R.T.: 4.216 min
Delta R.T.: -0.048 min
Response: 15041939
Conc: 3.24 ng/ml



#6 beta-BHC


R.T.: 0.000 min
Exp R.T.: 4.726 min
Response: 0
Conc: N.D.



#8 Heptachlor epoxide


R.T.: 0.000 min
Exp R.T.: 4.966 min
Response: 0
Conc: N.D.

Instrument :
ECD_L
ClientSampleId :




#8 Heptachlor epoxide

R.T.: 5.774 min
Delta R.T.: 0.024 min
Response: 7382127
Conc: 2.20 ng/ml




#10 gamma-Chlordane

R.T.: 5.205 min
Delta R.T.: 0.018 min
Response: 5803651
Conc: 0.59 ng/ml



#10 gamma-Chlordane


R.T.: 0.000 min
Exp R.T.: 5.983 min
Response: 0
Conc: N.D.



#11 alpha-Chlordane


R.T.: 5.205 min
Delta R.T.: -0.039 min
Response: 5803651
Conc: 0.60 ng/ml

Instrument :
ECD_L
ClientSampleId :




#11 alpha-Chlordane

R.T.: 0.000 min
Exp R.T. : 6.056 min
Response: 0
Conc: N.D.




#17 4,4'-DDT

R.T.: 6.151 min
Delta R.T.: -0.005 min
Response: 5044372
Conc: 0.67 ng/ml



#17 4,4'-DDT


R.T.: 0.000 min
Exp R.T. : 6.991 min
Response: 0
Conc: N.D.



#25 Chlordane-3


R.T.: 5.205 min
Delta R.T.: 0.017 min
Response: 5803651
Conc: 5.31 ng/ml

Instrument :
ECD_L
ClientSampleId :




#25 Chlordane-3

R.T.: 0.000 min
Exp R.T. : 5.983 min
Response: 0
Conc: N.D.




#26 Chlordane-4

R.T.: 5.205 min
Delta R.T.: -0.038 min
Response: 5803651
Conc: 5.81 ng/ml



#26 Chlordane-4

R.T.: 0.000 min
Exp R.T. : 6.058 min
Response: 0
Conc: N.D.



#28 Decachlorobiphenyl


R.T.: 7.963 min

Delta R.T.: -0.002 min

Response: 76655230

Conc: 10.91 ng/ml

Instrument :
ECD_L
ClientSampleId :



#28 Decachlorobiphenyl

R.T.: 8.895 min

Delta R.T.: 0.000 min

Response: 30902534

Conc: 13.32 ng/ml