

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL081820\
 Data File : PL060811.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Aug 2020 18:33
 Operator : DD\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 19 06:03:13 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL081820.M
 Quant Title : GC Extractables
 QLast Update : Wed Aug 19 02:25:28 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.462	4.010	80387661	135.1E6	48.153	48.577
28)	SA Decachlor...	8.219	9.100	69451241	100.6E6	48.620	49.781
Target Compounds							
2)	A alpha-BHC	3.899	4.400	120.2E6	212.4E6	49.380	49.994
3)	MA gamma-BHC...	4.182	4.685	108.5E6	195.7E6	49.998	49.634
4)	MA Heptachlor	4.484	5.196	99612031	188.4E6	50.183	50.125
5)	MB Aldrin	4.732	5.502	88135675	171.6E6	50.535	49.859
6)	B beta-BHC	4.433	4.856	44973096	80575481	50.060	49.870
7)	B delta-BHC	4.638	5.073	102.4E6	150.4E6	50.362	51.042
8)	B Heptachlo...	5.175	5.890	84643691	135.5E6	50.212	48.487
9)	A Endosulfan I	5.514	6.248	77854351	141.3E6	50.447	47.774
10)	B gamma-Chl...	5.403	6.125	83323851	149.0E6	50.645	48.798
11)	B alpha-Chl...	5.461	6.198	83306917	150.7E6	50.591	47.806
12)	B 4,4'-DDE	5.627	6.355	76308135	141.6E6	49.485	46.771
13)	MA Dieldrin	5.757	6.505	81177178	148.1E6	50.499	48.136
14)	MA Endrin	6.014	6.724	76004980	125.6E6	50.238	48.338
15)	B Endosulfa...	6.287	6.934	75585913	127.9E6	50.125	49.034
16)	A 4,4'-DDD	6.141	6.847	65293508	115.0E6	50.964	49.306
17)	MA 4,4'-DDT	6.382	7.147	63803670	114.6E6	53.207	52.795
18)	B Endrin al...	6.455	7.058	64613146	108.7E6	50.102	50.221
19)	B Endosulfa...	6.669	7.282	72246402	121.4E6	49.812	51.177
20)	A Methoxychlor	6.929	7.607	39105724	63436029	51.665	50.695
21)	B Endrin ke...	7.158	7.754	81041671	117.9E6	52.174	51.154
22)	Mirex	7.348	8.210	62314213	83392987	52.050	51.624

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL081820\
Data File : PL060811.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Aug 2020 18:33
Operator : DD\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 19 06:03:13 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL081820.M
Quant Title : GC Extractables
QLast Update : Wed Aug 19 02:25:28 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

