

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\
 Method File : LTX082021.M
 Title : GC Extractables
 Last Update : Fri Aug 20 06:11:45 2021
 Response Via : Initial Calibration

Calibration Files

500 =PL069254.D 1000 =PL069252.D 250 =PL069255.D
 750 =PL069253.D 100 =PL069256.D

	Compound	500	1000	250	750	100	Avg		%RSD

1) SA	Tetrachloro-m-xylene	4.557	4.679	4.516	4.617	4.802	4.634	E5	2.43
2)	Toxaphene-1	4.216	4.043	4.230	4.193	4.482	4.233	E3	3.73
3)	Toxaphene-2	1.085	1.092	1.115	1.080	1.223	1.119	E4	5.32
4)	Toxaphene-3	5.597	5.420	5.599	5.432	5.501	5.510	E3	1.56
5)	Toxaphene-4	0.957	0.950	0.975	0.942	1.032	0.971	E4	3.71
6)	Toxaphene-5	1.304	1.371	1.296	1.339	1.286	1.319	E4	2.67
7) SA	Decachlorobiphenyl	6.407	6.265	6.818	6.392	7.732	6.723	E5	8.94

Signal #2 Calibration Files

500 =PL069254.D 1000 =PL069252.D 250 =PL069255.D
 750 =PL069253.D 100 =PL069256.D

	Compound	500	1000	250	750	100	Avg		%RSD

1) SA	Tetrachloro-m-xylene	1.425	1.482	1.387	1.449	1.468	1.442	E6	2.61
2)	Toxaphene-1	7.776	8.020	8.097	8.088	7.621	7.920	E3	2.68
3)	Toxaphene-2	1.254	1.244	1.228	1.254	1.159	1.228	E4	3.25
4)	Toxaphene-3	1.044	1.079	1.084	1.033	1.185	1.085	E4	5.52
5)	Toxaphene-4	1.983	2.011	2.021	1.954	2.061	2.006	E4	2.01
6)	Toxaphene-5	1.526	1.527	1.546	1.474	1.550	1.524	E4	1.99
7) SA	Decachlorobiphenyl	1.189	1.193	1.222	1.194	1.340	1.227	E6	5.23

(#) = Out of Range