

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL082118\
 Data File : PL039396.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Aug 2018 03:58
 Operator : AJ\SJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 22 05:34:13 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL081318.M
 Quant Title : GC Extractables
 QLast Update : Wed Aug 15 15:04:22 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.305	3.907	352.2E6	152.7E6	50.384	56.970
2)	SA Decachlor...	7.962	8.897	298.3E6	117.1E6	42.468	50.441
Target Compounds							
2)	A alpha-BHC	3.732	4.285	601.3E6	241.9E6	50.869	57.443
3)	MA gamma-BHC...	4.008	4.564	546.4E6	216.7E6	51.515	53.811
4)	MA Heptachlor	4.290	5.071	527.8E6	212.4E6	50.135	59.386
5)	MB Aldrin	4.527	5.372	514.3E6	200.2E6	51.351	56.552
6)	B beta-BHC	4.263	4.726	236.1E6	98676412	50.818	56.337
7)	B delta-BHC	4.456	4.939	554.6E6	217.9E6	51.090	54.720
8)	B Heptachlo...	4.964	5.751	481.6E6	180.2E6	50.436	53.584
9)	A Endosulfan I	5.294	6.103	437.0E6	152.9E6	49.908	52.123
10)	B gamma-Chl...	5.185	5.983	491.7E6	175.3E6	49.780	54.151
11)	B alpha-Chl...	5.241	6.056	471.1E6	170.6E6	49.036	52.724
12)	B 4,4'-DDE	5.410	6.213	473.2E6	163.2E6	49.412	52.777
13)	MA Dieldrin	5.532	6.356	480.4E6	157.0E6	49.265	51.821
14)	MA Endrin	5.784	6.570	409.5E6	125.4E6	45.754	48.772
15)	B Endosulfa...	6.060	6.774	401.4E6	131.2E6	46.157	51.060
16)	A 4,4'-DDD	5.920	6.694	375.5E6	119.5E6	47.122	53.384
17)	MA 4,4'-DDT	6.154	6.992	341.8E6	110.1E6	45.089	46.777
18)	B Endrin al...	6.227	6.896	313.3E6	109.8E6	46.235	50.573
19)	B Endosulfa...	6.438	7.118	369.7E6	121.9E6	45.430	50.832
20)	A Methoxychlor	6.701	7.448	181.0E6	60386471	44.231	48.756
21)	B Endrin ke...	6.920	7.581	389.8E6	123.9E6	45.288	50.159

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL082118\
Data File : PL039396.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Aug 2018 03:58
Operator : AJ\SJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 22 05:34:13 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL081318.M
Quant Title : GC Extractables
QLast Update : Wed Aug 15 15:04:22 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

