

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL082222\
 Data File : PL077265.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Aug 2022 22:31
 Operator : AR\AJ
 Sample : N4260-04
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 SASP2-T-4-(16-22)

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 23 04:22:09 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL082222.M
 Quant Title : GC Extractables
 QLast Update : Mon Aug 22 19:19:43 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.415	2.630	23873056	54842376	17.545	17.845
28)	SA Decachlor...	8.904	7.734	18568474	43990816	15.845	15.765

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL082222\
Data File : PL077265.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Aug 2022 22:31
Operator : AR\AJ
Sample : N4260-04
Misc :
ALS Vial : 27 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
SASP2-T-4-(16-22)

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 23 04:22:09 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL082222.M
Quant Title : GC Extractables
QLast Update : Mon Aug 22 19:19:43 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene


R.T.: 3.415 min
Delta R.T.: -0.002 min
Response: 23873056
Conc: 17.55 ng/ml

Instrument :
ECD_L
ClientSampleId :
SASP2-T-4-(16-22)




#1 Tetrachloro-m-xylene

R.T.: 2.630 min
Delta R.T.: 0.000 min
Response: 54842376
Conc: 17.84 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.904 min
Delta R.T.: -0.003 min
Response: 18568474
Conc: 15.84 ng/ml



#28 Decachlorobiphenyl

R.T.: 7.734 min
Delta R.T.: -0.001 min
Response: 43990816
Conc: 15.77 ng/ml