

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL082323\
Data File : PL084956.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Aug 2023 21:45
Operator : AR\AJ
Sample : 04105-20MS
Misc :
ALS Vial : 30 Sample Multiplier: 1


Instrument :
ECD_L
ClientSampleId :
BH924MS

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 08/25/2023
Supervised By :mohammad ahmed 08/25/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 23 22:38:02 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL081523CLP.M
Quant Title : GC Extractables
QLast Update : Tue Aug 15 05:52:51 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm



(1) Tetrachloro-m-xylene (SA)

3.477min 19.865 ng/ml

response 49799667

(1) Tetrachloro-m-xylene #2 (SA)

2.685min 19.051 ng/ml

response 22490214

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL082323\
Data File : PL084956.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Aug 2023 21:45
Operator : AR\AJ
Sample : 04105-20MS
Misc :
ALS Vial : 30 Sample Multiplier: 1


Instrument :
ECD_L
ClientSampleId :
BH924MS

Manual IntegrationsAPPROVED

Reviewed By :Abdul Mirza 08/25/2023
Supervised By :mohammad ahmed 08/25/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 23 22:38:02 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL081523CLP.M
Quant Title : GC Extractables
QLast Update : Tue Aug 15 05:52:51 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm



(1) Tetrachloro-m-xylene (SA)

3.475min 19.017 ng/ml m

response 47673272

(1) Tetrachloro-m-xylene #2 (SA)

2.685min 19.051 ng/ml

response 22490214