

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL082520\
 Data File : PL060954.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 25 Aug 2020 15:43
 Operator : DD\AJ
 Sample : PSTDICC075
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDICC075

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 25 16:21:23 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL082520.M
 Quant Title : GC Extractables
 QLast Update : Tue Aug 25 16:20:29 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.461	4.010	138.0E6	196.7E6	75.122	75.553
28)	SA Decachlor...	8.217	9.099	122.1E6	153.8E6	74.180	74.568
Target Compounds							
2)	A alpha-BHC	3.898	4.399	217.2E6	316.6E6	77.092	77.176
3)	MA gamma-BHC...	4.181	4.685	194.4E6	291.6E6	76.519	76.828
4)	MA Heptachlor	4.483	5.196	178.0E6	279.0E6	75.867	76.039
5)	MB Aldrin	4.731	5.502	159.8E6	259.5E6	76.914	76.548
6)	B beta-BHC	4.432	4.856	75911976	116.2E6	75.438	75.634
7)	B delta-BHC	4.636	5.074	184.8E6	227.4E6	76.966	77.137
8)	B Heptachlo...	5.174	5.889	148.8E6	206.8E6	76.022	75.625
9)	A Endosulfan I	5.513	6.247	137.1E6	215.8E6	76.002	75.614
10)	B gamma-Chl...	5.402	6.125	149.3E6	232.5E6	77.021	75.096
11)	B alpha-Chl...	5.460	6.198	147.8E6	231.3E6	76.279	74.898
12)	B 4,4'-DDE	5.625	6.355	142.2E6	222.9E6	77.172	76.930
13)	MA Dieldrin	5.756	6.505	147.3E6	228.9E6	76.506	77.051
14)	MA Endrin	6.013	6.724	136.1E6	194.1E6	75.617	76.230
15)	B Endosulfa...	6.286	6.934	132.4E6	193.5E6	75.560	75.539
16)	A 4,4'-DDD	6.141	6.847	120.1E6	179.8E6	76.666	77.034
17)	MA 4,4'-DDT	6.382	7.147	122.1E6	181.4E6	77.893	77.557
18)	B Endrin al...	6.454	7.058	112.1E6	164.1E6	74.811	75.022
19)	B Endosulfa...	6.667	7.282	127.9E6	182.9E6	75.172	75.487
20)	A Methoxychlor	6.928	7.607	68075566	97471660	75.480	75.643
21)	B Endrin ke...	7.157	7.754	146.9E6	190.2E6	75.766	76.330
22)	Mirex	7.346	8.211	103.5E6	128.7E6	74.356	74.783

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL082520\
Data File : PL060954.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 25 Aug 2020 15:43
Operator : DD\AJ
Sample : PSTDICC075
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDICC075

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 25 16:21:23 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL082520.M
Quant Title : GC Extractables
QLast Update : Tue Aug 25 16:20:29 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

