

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL082521\
 Data File : PL0693442.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 25 Aug 2021 13:49
 Operator : AR\AJ
 Sample : M3502-04
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 26 03:48:05 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL082021.M
 Quant Title : GC Extractables
 QLast Update : Sat Aug 21 02:23:37 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.634	5.569	7428322	23972472	16.100	16.735
28) SA Decachlor...	10.402	11.493	10530372	20458075	15.602	16.693
Target Compounds						
7) B delta-BHC	0.000	7.044f	0	5895989	N.D.	3.192 #
9) A Endosulfan I	0.000	8.446	0	115025	N.D.	0.083 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL082521\
Data File : PL0693442.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 25 Aug 2021 13:49
Operator : AR\AJ
Sample : M3502-04
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampled :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 26 03:48:05 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL082021.M
Quant Title : GC Extractables
QLast Update : Sat Aug 21 02:23:37 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene


R.T.: 4.634 min
Delta R.T.: 0.000 min
Response: 7428322
Conc: 16.10 ng/ml

Instrument :
ECD_L
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 5.569 min
Delta R.T.: 0.002 min
Response: 23972472
Conc: 16.73 ng/ml




#7 delta-BHC

R.T.: 0.000 min
Exp R.T. : 6.404 min
Response: 0
Conc: N.D.



#7 delta-BHC


R.T.: 7.044 min
Delta R.T.: -0.020 min
Response: 5895989
Conc: 3.19 ng/ml



#9 Endosulfan I


R.T.: 0.000 min
Exp R.T.: 7.478 min
Response: 0
Conc: N.D.

Instrument :
ECD_L
ClientSampleId :




#9 Endosulfan I

R.T.: 8.446 min
Delta R.T.: 0.010 min
Response: 115025
Conc: 0.08 ng/ml



#28 Decachlorobiphenyl

R.T.: 10.402 min
Delta R.T.: 0.001 min
Response: 10530372
Conc: 15.60 ng/ml



#28 Decachlorobiphenyl

R.T.: 11.493 min
Delta R.T.: 0.005 min
Response: 20458075
Conc: 16.69 ng/ml