

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL082525\
 Data File : PL096960.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 25 Aug 2025 11:14
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 26 05:50:55 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL082025.M
 Quant Title : GC Extractables
 QLast Update : Thu Aug 21 06:34:01 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.530	2.825	189.7E6	279.2E6	48.098	49.531
28)	SA Decachlor...	9.007	7.987	137.3E6	250.2E6	46.371	52.166
Target Compounds							
2)	A alpha-BHC	3.977	3.331	293.6E6	433.8E6	50.665	52.828
3)	MA gamma-BHC...	4.305	3.663	275.7E6	403.1E6	50.471	54.716
4)	MA Heptachlor	4.897	4.011	270.0E6	398.0E6	50.634	52.891
5)	MB Aldrin	5.236	4.293	262.4E6	376.8E6	49.943	53.571
6)	B beta-BHC	4.492	3.958	113.1E6	172.0E6	49.434	52.737
7)	B delta-BHC	4.738	4.192	253.4E6	395.6E6	56.314	53.830
8)	B Heptachlo...	5.656	4.795	235.9E6	341.2E6	45.839	53.939
9)	A Endosulfan I	6.037	5.166	212.7E6	315.5E6	49.493	52.667
10)	B gamma-Chl...	5.909	5.047	232.3E6	353.7E6	50.437	53.508
11)	B alpha-Chl...	5.990	5.111	230.8E6	347.1E6	49.041	53.154
12)	B 4,4'-DDE	6.159	5.300	194.1E6	330.1E6	48.646	53.639
13)	MA Dieldrin	6.309	5.431	225.8E6	351.6E6	49.734	53.652
14)	MA Endrin	6.535	5.705	185.4E6	313.9E6	49.506	53.130
15)	B Endosulfa...	6.749	5.997	199.9E6	306.1E6	45.390	53.527
16)	A 4,4'-DDD	6.668	5.852	155.0E6	279.4E6	49.109	53.492
17)	MA 4,4'-DDT	6.983	6.104	171.9E6	297.0E6	49.086	54.221
18)	B Endrin al...	6.877	6.175	134.1E6	228.2E6	51.566	55.371
19)	B Endosulfa...	7.111	6.398	170.7E6	296.4E6	49.708	52.907
20)	A Methoxychlor	7.455	6.677	84326913	159.8E6	48.590	53.258
21)	B Endrin ke...	7.591	6.903	180.6E6	334.3E6	50.792	53.406
22)	Mirex	8.069	7.092	144.9E6	256.6E6	48.441	52.845

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL082525\
Data File : PL096960.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 25 Aug 2025 11:14
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 26 05:50:55 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL082025.M
Quant Title : GC Extractables
QLast Update : Thu Aug 21 06:34:01 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

