

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL082622\
 Data File : PL077337.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 26 Aug 2022 14:25
 Operator : AR\AJ
 Sample : N4378-01
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 CL-01-082522

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 26 22:30:06 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL082222.M
 Quant Title : GC Extractables
 QLast Update : Tue Aug 23 05:36:06 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.416	2.630	13862755	32335599	10.188	10.521
28)	SA Decachlor...	8.905	7.733	14797833	30370022	12.627	10.884
Target Compounds							
10)	B gamma-Chl...	5.819	4.800	3350478	3362364	2.389m	1.006m#
11)	B alpha-Chl...	5.899	4.863	6643039	7754202	4.722	2.369m#



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL082622\
Data File : PL077337.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 26 Aug 2022 14:25
Operator : AR\AJ
Sample : N4378-01
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
CL-01-082522

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 26 22:30:06 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL082222.M
Quant Title : GC Extractables
QLast Update : Tue Aug 23 05:36:06 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene


R.T.: 3.416 min
Delta R.T.: 0.000 min
Response: 13862755
Conc: 10.19 ng/ml

Instrument :
ECD_L
ClientSampleId :
CL-01-082522




#1 Tetrachloro-m-xylene

R.T.: 2.630 min
Delta R.T.: 0.000 min
Response: 32335599
Conc: 10.52 ng/ml




#10 gamma-Chlordane

R.T.: 5.819 min
Delta R.T.: 0.000 min
Response: 3350478
Conc: 2.39 ng/ml m



#10 gamma-Chlordane


R.T.: 4.800 min
Delta R.T.: -0.003 min
Response: 3362364
Conc: 1.01 ng/ml m



#11 alpha-Chlordane


R.T.: 5.899 min
Delta R.T.: 0.000 min
Response: 6643039
Conc: 4.72 ng/ml

Instrument :
ECD_L
ClientSampleId :
CL-01-082522




#11 alpha-Chlordane

R.T.: 4.863 min
Delta R.T.: -0.003 min
Response: 7754202
Conc: 2.37 ng/ml m



#28 Decachlorobiphenyl

R.T.: 8.905 min
Delta R.T.: 0.000 min
Response: 14797833
Conc: 12.63 ng/ml



#28 Decachlorobiphenyl

R.T.: 7.733 min
Delta R.T.: -0.001 min
Response: 30370022
Conc: 10.88 ng/ml