

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL082719\
 Data File : PL051863.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 28 Aug 2019 03:23
 Operator : SG\AJ
 Sample : K4517-15MS
 Misc :
 ALS Vial : 51 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 T4-S-CMS

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 28 03:55:51 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL081919.M
 Quant Title : GC Extractables
 QLast Update : Mon Aug 19 18:03:07 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.314	3.925	177.0E6	45899178	20.680	21.008
2)	SA Decachlor...	7.971	8.952	200.1E6	50013232	16.483	16.431
Target Compounds							
2)	A alpha-BHC	3.741	4.313	774.2E6	177.9E6	53.433	55.222
3)	MA gamma-BHC...	4.016	4.596	693.5E6	180.5E6	52.714	57.325
4)	MA Heptachlor	4.300	5.097	734.1E6	183.8E6	48.490	52.419
5)	MB Aldrin	4.537	5.399	677.1E6	178.3E6	46.853	54.389
6)	B beta-BHC	4.272	4.773	271.4E6	78073457	47.532	53.570
7)	B delta-BHC	4.464	4.987	670.3E6	157.2E6	45.083	50.214
8)	B Heptachlo...	4.973	5.785	632.8E6	169.4E6	45.247	50.191
9)	A Endosulfan I	5.303	6.141	625.2E6	159.5E6	46.505	49.321
10)	B gamma-Chl...	5.195	6.020	663.5E6	170.5E6	44.436	46.706
11)	B alpha-Chl...	5.252	6.090	642.8E6	169.0E6	44.416	48.984
12)	B 4,4'-DDE	5.418	6.251	621.7E6	162.9E6	45.428	46.271
13)	MA Dieldrin	5.542	6.397	716.6E6	177.9E6	47.127	51.955
14)	MA Endrin	5.794	6.614	646.8E6	158.9E6	44.343	52.535
15)	B Endosulfa...	6.069	6.826	579.2E6	165.4E6	41.626	50.186
16)	A 4,4'-DDD	5.929	6.742	507.7E6	150.5E6	41.460	55.607 #
17)	MA 4,4'-DDT	6.163	7.038	544.2E6	141.5E6	42.676	48.316
18)	B Endrin al...	6.236	6.950	459.3E6	132.9E6	39.479	47.024
19)	B Endosulfa...	6.448	7.173	590.6E6	161.4E6	45.398	52.276
20)	A Methoxychlor	6.707	7.499	287.9E6	78309385	42.467	46.691
21)	B Endrin ke...	6.928	7.642	633.0E6	174.1E6	44.742	54.411
22)	Mirex	7.102	8.089	424.5E6	113.7E6	40.098	46.736

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL082719\
Data File : PL051863.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 28 Aug 2019 03:23
Operator : SG\AJ
Sample : K4517-15MS
Misc :
ALS Vial : 51 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampled :
T4-S-CMS

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 28 03:55:51 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL081919.M
Quant Title : GC Extractables
QLast Update : Mon Aug 19 18:03:07 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

