

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL082719\
 Data File : PL051886.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 28 Aug 2019 10:33
 Operator : SG\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Manual Integrations
 APPROVED

Ankita
 8/28/2019 5:04:34 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 28 10:46:31 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL081919.M
 Quant Title : GC Extractables
 QLast Update : Mon Aug 19 18:03:07 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.312	3.925	370.5E6	73611136	43.295	33.692
2)	SA Decachlor...	7.969	8.950	516.9E6	131.0E6	42.589	43.042
Target Compounds							
2)	A alpha-BHC	3.739	4.313	637.8E6	115.6E6	44.019	35.860
3)	MA gamma-BHC...	4.014	4.596	584.2E6	115.1E6	44.406	36.557
4)	MA Heptachlor	4.297	5.096	673.0E6	135.2E6	44.458	38.540
5)	MB Aldrin	4.535	5.398	630.1E6	132.2E6	43.604	40.308
6)	B beta-BHC	4.269	4.773	245.1E6	56124158	42.931	38.509
7)	B delta-BHC	4.461	4.986	649.8E6	119.7E6	43.704	38.247
8)	B Heptachlo...	4.970	5.785	625.7E6	127.6E6	44.739	37.798
9)	A Endosulfan I	5.300	6.139	632.8E6	130.4E6	47.067	40.332
10)	B gamma-Chl...	5.192	6.018	678.0E6	139.1E6	45.406	38.099
11)	B alpha-Chl...	5.248	6.090	661.5E6	140.0E6	45.713	40.579
12)	B 4,4'-DDE	5.415	6.250	700.0E6	149.7E6	51.152	42.517
13)	MA Dieldrin	5.539	6.396	700.6E6	140.8E6	46.074	41.113
14)	MA Endrin	5.791	6.613	671.9E6	120.7E6	46.066	39.909
15)	B Endosulfa...	6.067	6.825	585.0E6	143.1E6	42.041	43.412
16)	A 4,4'-DDD	5.926	6.741	525.4E6	121.3E6	42.904	44.820
17)	MA 4,4'-DDT	6.159	7.036	587.1E6	131.8E6	46.045	45.023m
18)	B Endrin al...	6.233	6.949	502.0E6	113.8E6	43.151	40.250
19)	B Endosulfa...	6.444	7.172	553.1E6	126.9E6	42.514	41.099
20)	A Methoxychlor	6.704	7.498	280.3E6	71435870	41.343	42.592
21)	B Endrin ke...	6.925	7.640	610.2E6	136.2E6	43.135	42.565
22)	Mirex	7.098	8.088	450.9E6	109.8E6	42.591	45.146

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL082719\
Data File : PL051886.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 28 Aug 2019 10:33
Operator : SG\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Manual Integrations
APPROVED

Ankita
8/28/2019 5:04:34 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 28 10:46:31 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL081919.M
Quant Title : GC Extractables
QLast Update : Mon Aug 19 18:03:07 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

