

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL082919\
 Data File : PL052020.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 30 Aug 2019 03:34
 Operator : SG\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Manual Integrations
 APPROVED

Ankita
 8/31/2019 2:40:31 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 30 08:05:02 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL081919.M
 Quant Title : GC Extractables
 QLast Update : Mon Aug 19 18:03:07 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.313	3.924	484.3E6	117.3E6	56.582	53.697
28)	SA Decachlor...	7.971	8.948	570.7E6	143.9E6	47.015	47.259
Target Compounds							
2)	A alpha-BHC	3.741	4.312	809.7E6	160.0E6	55.883	49.640
3)	MA gamma-BHC...	4.016	4.595	752.8E6	155.7E6	57.217	49.442
4)	MA Heptachlor	4.299	5.096	801.4E6	175.3E6	52.934	49.996
5)	MB Aldrin	4.537	5.398	783.1E6	170.3E6	54.194	51.935
6)	B beta-BHC	4.271	4.772	314.7E6	75782713	55.118	51.998
7)	B delta-BHC	4.463	4.986	814.0E6	168.4E6	54.746	53.819
8)	B Heptachlo...	4.973	5.784	749.7E6	171.7E6	53.606	50.864
9)	A Endosulfan I	5.303	6.139	720.8E6	162.9E6	53.614	50.397
10)	B gamma-Chl...	5.195	6.018	809.8E6	176.1E6	54.230	48.246
11)	B alpha-Chl...	5.250	6.089	777.7E6	174.6E6	53.744	50.603
12)	B 4,4'-DDE	5.418	6.250	714.2E6	180.6E6	52.189	51.316
13)	MA Dieldrin	5.541	6.396	783.4E6	175.5E6	51.524	51.256
14)	MA Endrin	5.794	6.612	693.0E6	148.2E6	47.510	49.006
15)	B Endosulfa...	6.069	6.825	677.3E6	159.6E6	48.674	48.409
16)	A 4,4'-DDD	5.928	6.741	628.0E6	142.7E6	51.286	52.725
17)	MA 4,4'-DDT	6.162	7.036	606.5E6	142.1E6	47.565	48.520m
18)	B Endrin al...	6.236	6.949	571.7E6	137.4E6	49.140	48.602
19)	B Endosulfa...	6.446	7.172	637.7E6	152.7E6	49.016m	49.434
20)	A Methoxychlor	6.706	7.498	314.2E6	77253115	46.350	46.061
21)	B Endrin ke...	6.928	7.640	703.1E6	161.6E6	49.697	50.497
22)	Mirex	7.101	8.087	483.2E6	118.6E6	45.638	48.756

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL082919\
Data File : PL052020.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 Aug 2019 03:34
Operator : SG\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Manual Integrations
APPROVED

Ankita
8/31/2019 2:40:31 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 30 08:05:02 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL081919.M
Quant Title : GC Extractables
QLast Update : Mon Aug 19 18:03:07 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

