

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL082922\
 Data File : PL077361.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 29 Aug 2022 16:18
 Operator : AR\AJ
 Sample : PB147331BL
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 29 18:13:26 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL082222.M
 Quant Title : GC Extractables
 QLast Update : Tue Aug 23 05:36:06 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.415	2.630	21408294	48693763	15.734	15.844
28)	SA Decachlor...	8.904	7.731	19741793	41693746	16.846	14.942
Target Compounds							
2)	A alpha-BHC	0.000	3.109	0	955303	N.D.	0.219 #
14)	MA Endrin	0.000	5.476f	0	143823	N.D.	0.053 #
15)	B Endosulfa...	6.707f	0.000	266537	0	0.243	N.D. #
24)	Chlordane-2	0.000	4.197f	0	1724860	N.D.	16.102 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL082922\
Data File : PL077361.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 29 Aug 2022 16:18
Operator : AR\AJ
Sample : PB147331BL
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 29 18:13:26 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL082222.M
Quant Title : GC Extractables
QLast Update : Tue Aug 23 05:36:06 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene


R.T.: 3.415 min
Delta R.T.: 0.000 min
Response: 21408294
Conc: 15.73 ng/ml

Instrument :
ECD_L
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 2.630 min
Delta R.T.: 0.000 min
Response: 48693763
Conc: 15.84 ng/ml




#2 alpha-BHC

R.T.: 0.000 min
Exp R.T. : 3.880 min
Response: 0
Conc: N.D.



#2 alpha-BHC


R.T.: 3.109 min
Delta R.T.: -0.014 min
Response: 955303
Conc: 0.22 ng/ml



#14 Endrin


R.T.: 0.000 min
Exp R.T.: 6.460 min
Response: 0
Conc: N.D.

Instrument :
ECD_L
ClientSampleId :




#14 Endrin

R.T.: 5.476 min
Delta R.T.: 0.020 min
Response: 143823
Conc: 0.05 ng/ml




#15 Endosulfan II

R.T.: 6.707 min
Delta R.T.: 0.019 min
Response: 266537
Conc: 0.24 ng/ml



#15 Endosulfan II


R.T.: 0.000 min
Exp R.T.: 5.752 min
Response: 0
Conc: N.D.



#24 Chlordane-2


R.T.: 0.000 min
Exp R.T.: 5.109 min
Response: 0
Conc: N.D.

Instrument :
ECD_L
ClientSampleId :




#24 Chlordane-2

R.T.: 4.197 min
Delta R.T.: 0.017 min
Response: 1724860
Conc: 16.10 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.904 min
Delta R.T.: -0.001 min
Response: 19741793
Conc: 16.85 ng/ml



#28 Decachlorobiphenyl

R.T.: 7.731 min
Delta R.T.: -0.004 min
Response: 41693746
Conc: 14.94 ng/ml