

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL082923\
 Data File : PL085041.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 29 Aug 2023 13:59
 Operator : AR\AJ
 Sample : 04196-15
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 OR-345-COMP-03

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 29 20:50:00 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL080323.M
 Quant Title : GC Extractables
 QLast Update : Thu Aug 03 16:21:43 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.477	2.685	24029135	10924478	10.342	10.046
28)	SA Decachlor...	9.029	7.865	18689053	13367462	11.486	10.721
Target Compounds							
12)	B 4,4'-DDE	6.150	5.162	151.4E6	89472301	65.971	72.655
16)	A 4,4'-DDD	6.670	5.719	3919655	1908011	2.338	1.847
17)	MA 4,4'-DDT	6.985	5.972	71988871	45694500	38.421	39.373



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL082923\
Data File : PL085041.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 29 Aug 2023 13:59
Operator : AR\AJ
Sample : 04196-15
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
OR-345-COMP-03

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 29 20:50:00 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL080323.M
Quant Title : GC Extractables
QLast Update : Thu Aug 03 16:21:43 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene


R.T.: 3.477 min
Delta R.T.: 0.001 min
Response: 24029135
Conc: 10.34 ng/ml

Instrument :
ECD_L
ClientSampleId :
OR-345-COMP-03




#1 Tetrachloro-m-xylene

R.T.: 2.685 min
Delta R.T.: -0.002 min
Response: 10924478
Conc: 10.05 ng/ml




#12 4,4'-DDE

R.T.: 6.150 min
Delta R.T.: 0.002 min
Response: 151421580
Conc: 65.97 ng/ml



#12 4,4'-DDE


R.T.: 5.162 min
Delta R.T.: -0.003 min
Response: 89472301
Conc: 72.66 ng/ml



#16 4,4'-DDD


R.T.: 6.670 min
Delta R.T.: 0.003 min
Response: 3919655
Conc: 2.34 ng/ml

Instrument :
ECD_L
ClientSampleId :
OR-345-COMP-03




#16 4,4'-DDD

R.T.: 5.719 min
Delta R.T.: -0.004 min
Response: 1908011
Conc: 1.85 ng/ml




#17 4,4'-DDT

R.T.: 6.985 min
Delta R.T.: 0.002 min
Response: 71988871
Conc: 38.42 ng/ml



#17 4,4'-DDT


R.T.: 5.972 min
Delta R.T.: -0.003 min
Response: 45694500
Conc: 39.37 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.029 min
Delta R.T.: 0.004 min
Response: 18689053
Conc: 11.49 ng/ml

Instrument :
ECD_L
ClientSampleId :
OR-345-COMP-03



#28 Decachlorobiphenyl

R.T.: 7.865 min
Delta R.T.: -0.003 min
Response: 13367462
Conc: 10.72 ng/ml