

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL083018\
 Data File : PL039613.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 30 Aug 2018 17:24
 Operator : AJ\SJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Manual Integrations
 APPROVED

Sohil
 8/31/2018 12:25:13 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 31 02:09:49 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL081318.M
 Quant Title : GC Extractables
 QLast Update : Wed Aug 15 15:04:22 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.303	3.906	342.7E6	146.4E6	49.023	54.603
28)	SA Decachlor...	7.957	8.892	301.4E6	115.8E6	42.919	49.899
Target Compounds							
2)	A alpha-BHC	3.730	4.283	582.3E6	234.2E6	49.267	55.613
3)	MA gamma-BHC...	4.006	4.562	528.4E6	209.6E6	49.814	52.048
4)	MA Heptachlor	4.287	5.069	501.5E6	204.0E6	47.642	57.062
5)	MB Aldrin	4.524	5.369	502.9E6	189.5E6	50.207	53.527
6)	B beta-BHC	4.261	4.724	226.2E6	95455560	48.689	54.498
7)	B delta-BHC	4.453	4.937	538.1E6	208.6E6	49.575	52.382
8)	B Heptachlo...	4.960	5.748	466.9E6	172.4E6	48.892	51.259
9)	A Endosulfan I	5.289	6.099	426.8E6	148.5E6	48.745	50.609
10)	B gamma-Chl...	5.182	5.980	477.7E6	168.2E6	48.362	51.984
11)	B alpha-Chl...	5.237	6.053	461.8E6	164.3E6	48.068	50.777
12)	B 4,4'-DDE	5.406	6.209	455.3E6	159.1E6	47.544	51.452
13)	MA Dieldrin	5.528	6.353	463.1E6	151.2E6	47.484	49.903
14)	MA Endrin	5.780	6.566	391.5E6	121.7E6	43.736	47.346
15)	B Endosulfa...	6.056	6.770	400.1E6	130.2E6	46.009	50.694
16)	A 4,4'-DDD	5.917	6.690	368.8E6	120.8E6	46.281	53.941
17)	MA 4,4'-DDT	6.149	6.989	325.1E6	106.8E6	42.881	45.375
18)	B Endrin al...	6.223	6.892	306.7E6	109.8E6	45.259	50.566
19)	B Endosulfa...	6.434	7.114	364.1E6	119.3E6	44.746	49.772
20)	A Methoxychlor	6.697	7.444	171.0E6	56749253	41.798	45.819
21)	B Endrin ke...	6.914	7.577	413.2E6	120.8E6	48.012m	48.877

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL083018\
Data File : PL039613.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 Aug 2018 17:24
Operator : AJ\SJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Manual Integrations
APPROVED

Sohil
8/31/2018 12:25:13 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 31 02:09:49 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL081318.M
Quant Title : GC Extractables
QLast Update : Wed Aug 15 15:04:22 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

