

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL083119\
 Data File : PL052150.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 31 Aug 2019 22:56
 Operator : SG\AJ
 Sample : PB122743BL
 Misc :
 ALS Vial : 38 Sample Multiplier: 1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 01 00:49:56 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL081919.M
 Quant Title : GC Extractables
 QLast Update : Mon Aug 19 18:03:07 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml



System Monitoring Compounds							
1)	SA Tetrachlo...	3.312	3.923	173.8E6	46268822	20.308	21.177
28)	SA Decachlor...	7.967	8.945	234.6E6	55526833	19.329	18.242
Target Compounds							
6)	B beta-BHC	0.000	4.799f	0	930872	N.D.	0.639 #
8)	B Heptachlo...	0.000	5.799	0	1364506	N.D.	0.404 #
15)	B Endosulfa...	0.000	6.841	0	2246710	N.D.	0.682 #
20)	A Methoxychlor	0.000	7.484f	0	88277	N.D.	0.053 #


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL083119\
Data File : PL052150.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 31 Aug 2019 22:56
Operator : SG\AJ
Sample : PB122743BL
Misc :
ALS Vial : 38 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 01 00:49:56 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL081919.M
Quant Title : GC Extractables
QLast Update : Mon Aug 19 18:03:07 2019
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene

R.T.: 3.312 min
Delta R.T.: -0.004 min
Response: 173808600
Conc: 20.31 ng/ml




#1 Tetrachloro-m-xylene

R.T.: 3.923 min
Delta R.T.: -0.004 min
Response: 46268822
Conc: 21.18 ng/ml




#6 beta-BHC

R.T.: 0.000 min
Exp R.T. : 4.273 min
Response: 0
Conc: N.D.




#6 beta-BHC

R.T.: 4.799 min
Delta R.T.: 0.023 min
Response: 930872
Conc: 0.64 ng/ml




#8 Heptachlor epoxide

R.T.: 0.000 min
Exp R.T. : 4.975 min
Response: 0
Conc: N.D.




#8 Heptachlor epoxide

R.T.: 5.799 min
Delta R.T.: 0.012 min
Response: 1364506
Conc: 0.40 ng/ml




#15 Endosulfan II

R.T.: 0.000 min
Exp R.T. : 6.072 min
Response: 0
Conc: N.D.




#15 Endosulfan II

R.T.: 6.841 min
Delta R.T.: 0.012 min
Response: 2246710
Conc: 0.68 ng/ml




#20 Methoxychlor

R.T.: 0.000 min
Exp R.T.: 6.710 min
Response: 0
Conc: N.D.




#20 Methoxychlor

R.T.: 7.484 min
Delta R.T.: -0.017 min
Response: 88277
Conc: 0.05 ng/ml



#28 Decachlorobiphenyl

R.T.: 7.967 min
Delta R.T.: -0.008 min
Response: 234611943
Conc: 19.33 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.945 min
Delta R.T.: -0.009 min
Response: 55526833
Conc: 18.24 ng/ml