

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL083119\
 Data File : PL052134.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 31 Aug 2019 19:10
 Operator : SG\AJ
 Sample : K4554-07MSD
 Misc :
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 SB-15-(0-2)MSD

Manual Integrations
 APPROVED

Ankita
 9/3/2019 2:15:48 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 31 22:43:44 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL081919.M
 Quant Title : GC Extractables
 QLast Update : Mon Aug 19 18:03:07 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.313	3.923	199.2E6	52880211	23.275	24.204
28)	SA Decachlor...	7.969	8.947	256.5E6	59750649	21.131	19.630
Target Compounds							
2)	A alpha-BHC	3.740	4.310	926.5E6	208.7E6	63.944	64.779m
3)	MA gamma-BHC...	4.015	4.594	873.1E6	199.7E6	66.363	63.429
4)	MA Heptachlor	4.298	5.094	956.3E6	215.6E6	63.166	61.480
5)	MB Aldrin	4.535	5.396	951.2E6	214.5E6	65.827	65.426
6)	B beta-BHC	4.270	4.771	365.3E6	92988063	63.968	63.803
7)	B delta-BHC	4.462	4.984	877.9E6	175.7E6	59.043	56.123
8)	B Heptachlor...	4.972	5.783	863.0E6	205.6E6	61.704	60.930
9)	A Endosulfan I	5.301	6.137	832.1E6	196.0E6	61.893	60.620
10)	B gamma-Chl...	5.193	6.016	916.0E6	207.6E6	61.342	56.873
11)	B alpha-Chl...	5.249	6.087	886.0E6	211.6E6	61.226	61.318
12)	B 4,4'-DDE	5.416	6.249	908.1E6	214.2E6	66.357	60.863
13)	MA Dieldrin	5.539	6.394	920.9E6	212.4E6	60.563	62.010
14)	MA Endrin	5.792	6.610	849.8E6	181.6E6	58.259	60.029
15)	B Endosulfa...	6.068	6.823	773.7E6	189.4E6	55.608	57.454
16)	A 4,4'-DDD	5.926	6.739	716.9E6	172.0E6	58.544	63.530
17)	MA 4,4'-DDT	6.160	7.035	745.2E6	166.3E6	58.439	56.797
18)	B Endrin al...	6.234	6.947	627.7E6	159.3E6	53.960	56.368
19)	B Endosulfa...	6.445	7.170	741.8E6	179.6E6	57.023	58.171
20)	A Methoxychlor	6.705	7.496	349.7E6	91916442	51.587	54.803
21)	B Endrin ke...	6.926	7.638	836.3E6	195.0E6	59.112	60.937
22)	Mirex	7.099	8.085	602.3E6	142.9E6	56.885	58.748

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL083119\
Data File : PL052134.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 31 Aug 2019 19:10
Operator : SG\AJ
Sample : K4554-07MSD
Misc :
ALS Vial : 25 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
SB-15-(0-2)MSD

Manual Integrations
APPROVED

Ankita
9/3/2019 2:15:48 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 31 22:43:44 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL081919.M
Quant Title : GC Extractables
QLast Update : Mon Aug 19 18:03:07 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

