

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL083121\
 Data File : PL069652.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 31 Aug 2021 23:16
 Operator : AR\AJ
 Sample : M3574-01
 Misc :
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 TPR-12-13-COMP

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 01 01:34:26 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL083021.M
 Quant Title : GC Extractables
 QLast Update : Tue Aug 31 01:49:30 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.633	5.568	7733437	28954019	24.841	23.014
28)	SA Decachlor...	10.401	11.491	12709493	24873329	21.551	22.042
Target Compounds							
9)	A Endosulfan I	7.502f	0.000	854686	0	1.808	N.D. #
13)	MA Dieldrin	7.772	0.000	963723	0	1.949	N.D. #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL083121\
Data File : PL069652.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 31 Aug 2021 23:16
Operator : AR\AJ
Sample : M3574-01
Misc :
ALS Vial : 30 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
TPR-12-13-COMP

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 01 01:34:26 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL083021.M
Quant Title : GC Extractables
QLast Update : Tue Aug 31 01:49:30 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene


R.T.: 4.633 min
Delta R.T.: -0.001 min
Response: 7733437
Conc: 24.84 ng/ml

Instrument :
ECD_L
ClientSampleId :
TPR-12-13-COMP




#1 Tetrachloro-m-xylene

R.T.: 5.568 min
Delta R.T.: -0.002 min
Response: 28954019
Conc: 23.01 ng/ml




#9 Endosulfan I

R.T.: 7.502 min
Delta R.T.: 0.024 min
Response: 854686
Conc: 1.81 ng/ml



#9 Endosulfan I


R.T.: 0.000 min
Exp R.T.: 8.439 min
Response: 0
Conc: N.D.



#13 Dieldrin


R.T.: 7.772 min
Delta R.T.: 0.007 min
Response: 963723
Conc: 1.95 ng/ml

Instrument :
ECD_L
ClientSampleId :
TPR-12-13-COMP




#13 Dieldrin

R.T.: 0.000 min
Exp R.T. : 8.729 min
Response: 0
Conc: N.D.



#28 Decachlorobiphenyl

R.T.: 10.401 min
Delta R.T.: -0.001 min
Response: 12709493
Conc: 21.55 ng/ml



#28 Decachlorobiphenyl

R.T.: 11.491 min
Delta R.T.: -0.002 min
Response: 24873329
Conc: 22.04 ng/ml