

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL083122\
 Data File : PL077420.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 31 Aug 2022 17:36
 Operator : AR\AJ
 Sample : PB147383BL
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 31 18:05:50 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL082222.M
 Quant Title : GC Extractables
 QLast Update : Tue Aug 23 05:36:06 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.416	2.630	23134185	49891199	17.002	16.234
28)	SA Decachlor...	8.907	7.733	20877821	41270439	17.816	14.791
Target Compounds							
7)	B delta-BHC	4.643f	0.000	330141	0	0.214	N.D. #
14)	MA Endrin	0.000	5.476f	0	391379	N.D.	0.143 #
15)	B Endosulfa...	6.709f	0.000	183791	0	0.168	N.D. #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL083122\
Data File : PL077420.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 31 Aug 2022 17:36
Operator : AR\AJ
Sample : PB147383BL
Misc :
ALS Vial : 24 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 31 18:05:50 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL082222.M
Quant Title : GC Extractables
QLast Update : Tue Aug 23 05:36:06 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene


R.T.: 3.416 min
Delta R.T.: 0.001 min
Response: 23134185
Conc: 17.00 ng/ml

Instrument :
ECD_L
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 2.630 min
Delta R.T.: 0.000 min
Response: 49891199
Conc: 16.23 ng/ml




#7 delta-BHC

R.T.: 4.643 min
Delta R.T.: -0.029 min
Response: 330141
Conc: 0.21 ng/ml



#7 delta-BHC


R.T.: 0.000 min
Exp R.T. : 3.970 min
Response: 0
Conc: N.D.



#14 Endrin


R.T.: 0.000 min
Exp R.T.: 6.460 min
Response: 0
Conc: N.D.

Instrument :
ECD_L
ClientSampleId :




#14 Endrin

R.T.: 5.476 min
Delta R.T.: 0.021 min
Response: 391379
Conc: 0.14 ng/ml




#15 Endosulfan II

R.T.: 6.709 min
Delta R.T.: 0.021 min
Response: 183791
Conc: 0.17 ng/ml



#15 Endosulfan II


R.T.: 0.000 min
Exp R.T.: 5.752 min
Response: 0
Conc: N.D.



#28 Decachlorobiphenyl

R.T.: 8.907 min
Delta R.T.: 0.002 min
Response: 20877821
Conc: 17.82 ng/ml

Instrument :
ECD_L
ClientSampleId :



#28 Decachlorobiphenyl

R.T.: 7.733 min
Delta R.T.: -0.001 min
Response: 41270439
Conc: 14.79 ng/ml