

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL083122\
 Data File : PL077431.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 31 Aug 2022 20:08
 Operator : AR\AJ
 Sample : N4438-03
 Misc :
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 KTS2-WC-S-MH-02-0-20

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 01 03:02:29 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL082222.M
 Quant Title : GC Extractables
 QLast Update : Tue Aug 23 05:36:06 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.417	2.630	23663376	49908098	17.391	16.239
28)	SA Decachlor...	8.906	7.731	16246713	32894796	13.864	11.789

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL083122\
Data File : PL077431.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 31 Aug 2022 20:08
Operator : AR\AJ
Sample : N4438-03
Misc :
ALS Vial : 35 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
KTS2-WC-S-MH-02-0-20

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 01 03:02:29 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL082222.M
Quant Title : GC Extractables
QLast Update : Tue Aug 23 05:36:06 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene


R.T.: 3.417 min
Delta R.T.: 0.002 min
Response: 23663376
Conc: 17.39 ng/ml

Instrument :
ECD_L
ClientSampleId :
KTS2-WC-S-MH-02-0-20




#1 Tetrachloro-m-xylene

R.T.: 2.630 min
Delta R.T.: 0.000 min
Response: 49908098
Conc: 16.24 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.906 min
Delta R.T.: 0.000 min
Response: 16246713
Conc: 13.86 ng/ml



#28 Decachlorobiphenyl

R.T.: 7.731 min
Delta R.T.: -0.003 min
Response: 32894796
Conc: 11.79 ng/ml