

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL090418\
 Data File : PL039692.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Sep 2018 15:32
 Operator : AJ\SJ
 Sample : J4756-02
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 WE-4

Manual Integrations
 APPROVED

Sohil
 9/5/2018 1:18:34 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 05 02:25:22 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL081318.M
 Quant Title : GC Extractables
 QLast Update : Wed Aug 15 15:04:22 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.302	3.905	94659299	37187130	13.541	13.871
28) SA Decachlor...	7.956	8.891	93991374	37907614	13.383	16.335
Target Compounds						
12) B 4,4'-DDE	5.405	6.206	168.1E6	56539126	17.553	18.286m
17) MA 4,4'-DDT	6.145	6.986	147.4E6	31572898	19.445m	13.413m#

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL090418\
Data File : PL039692.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Sep 2018 15:32
Operator : AJ\SJ
Sample : J4756-02
Misc :
ALS Vial : 9 Sample Multiplier: 1



Instrument :
ECD_L
ClientSampled :
WE-4


Manual Integrations
APPROVED

Sohil
9/5/2018 1:18:34 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 05 02:25:22 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL081318.M
Quant Title : GC Extractables
QLast Update : Wed Aug 15 15:04:22 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene

R.T.: 3.302 min
Delta R.T.: -0.004 min
Response: 94659299
Conc: 13.54 ng/ml

Instrument :
ECD_L
ClientSampleId :
WE-4


Manual Integrations
APPROVED

Sohil
9/5/2018 1:18:34 PM




#1 Tetrachloro-m-xylene

R.T.: 3.905 min
Delta R.T.: -0.002 min
Response: 37187130
Conc: 13.87 ng/ml




#12 4,4'-DDE

R.T.: 5.405 min
Delta R.T.: -0.007 min
Response: 168087981
Conc: 17.55 ng/ml



#12 4,4'-DDE

R.T.: 6.206 min
Delta R.T.: -0.006 min
Response: 56539126
Conc: 18.29 ng/ml m




#17 4,4'-DDT

R.T.: 6.145 min
Delta R.T.: -0.011 min
Response: 147401435
Conc: 19.44 ng/ml m

Instrument :
ECD_L
ClientSampleId :
WE-4


Manual Integrations
APPROVED

Sohil
9/5/2018 1:18:34 PM




#17 4,4'-DDT

R.T.: 6.986 min
Delta R.T.: -0.006 min
Response: 31572898
Conc: 13.41 ng/ml m



#28 Decachlorobiphenyl

R.T.: 7.956 min
Delta R.T.: -0.009 min
Response: 93991374
Conc: 13.38 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.891 min
Delta R.T.: -0.004 min
Response: 37907614
Conc: 16.34 ng/ml