

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL090519\
 Data File : PL052201.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 05 Sep 2019 09:37
 Operator : SG\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Manual Integrations
 APPROVED

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 06 00:47:55 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL081919.M
 Quant Title : GC Extractables
 QLast Update : Mon Aug 19 18:03:07 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.312	3.923	449.8E6	110.4E6	52.557	50.510
28)	SA Decachlor...	7.969	8.947	657.8E6	143.7E6	54.191	47.213
Target Compounds							
2)	A alpha-BHC	3.739	4.311	742.7E6	158.4E6	51.256	49.168
3)	MA gamma-BHC...	4.014	4.594	707.5E6	152.4E6	53.772	48.420
4)	MA Heptachlor	4.297	5.094	777.0E6	167.3E6	51.328	47.702
5)	MB Aldrin	4.535	5.396	736.7E6	162.1E6	50.982	49.435
6)	B beta-BHC	4.270	4.771	291.5E6	73030549	51.049	50.109
7)	B delta-BHC	4.462	4.984	772.1E6	155.6E6	51.927	49.732
8)	B Heptachlor...	4.971	5.782	697.8E6	160.8E6	49.895	47.658
9)	A Endosulfan I	5.301	6.137	716.3E6	157.9E6	53.285	48.845
10)	B gamma-Chl...	5.193	6.016	773.0E6	174.8E6	51.767	47.873
11)	B alpha-Chl...	5.248	6.087	756.2E6	169.2E6	52.254	49.024
12)	B 4,4'-DDE	5.416	6.247	742.9E6	176.4E6	54.284	50.102
13)	MA Dieldrin	5.539	6.393	781.9E6	171.9E6	51.426	50.187
14)	MA Endrin	5.792	6.611	750.6E6	154.6E6	51.457	51.110
15)	B Endosulfa...	6.067	6.823	752.0E6	158.7E6	54.047	48.149
16)	A 4,4'-DDD	5.926	6.739	620.1E6	139.1E6	50.637	51.363
17)	MA 4,4'-DDT	6.160	7.034	680.7E6	153.0E6	53.380	52.273m
18)	B Endrin al...	6.234	6.947	587.6E6	133.8E6	50.513	47.357
19)	B Endosulfa...	6.445	7.170	661.6E6	150.4E6	50.860	48.705
20)	A Methoxychlor	6.704	7.496	348.3E6	81386690	51.370	48.525
21)	B Endrin ke...	6.926	7.638	741.3E6	153.9E6	52.398	48.089
22)	Mirex	7.098	8.085	553.9E6	129.7E6	52.318	53.306

Ankita

9/6/2019 1:59:14 PM

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL090519\
Data File : PL052201.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 05 Sep 2019 09:37
Operator : SG\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Manual Integrations
APPROVED

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 06 00:47:55 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL081919.M
Quant Title : GC Extractables
QLast Update : Mon Aug 19 18:03:07 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

