

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL090618\
 Data File : PL039722.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Sep 2018 14:57
 Operator : AJ\SJ
 Sample : J4781-01
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 07 04:09:42 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL081318.M
 Quant Title : GC Extractables
 QLast Update : Wed Aug 15 15:04:22 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.302	3.905	182.1E6	79647465	26.048	29.709
28)	SA Decachlor...	7.955	8.891	168.5E6	66495723	23.994	28.654
Target Compounds							
5)	MB Aldrin	4.513	0.000	6607702	0	0.660	N.D. #
6)	B beta-BHC	4.219f	0.000	15856648	0	3.413	N.D. #
8)	B Heptachlo...	5.025f	0.000	20413355	0	2.138	N.D. #
14)	MA Endrin	5.772	0.000	8764928	0	0.979	N.D. #
15)	B Endosulfa...	6.077	6.797	17047224	3961138	1.961	1.542
16)	A 4,4'-DDD	5.929	0.000	6995758	0	0.878	N.D. #
20)	A Methoxychlor	6.704	0.000	6922135	0	1.692	N.D. #
21)	B Endrin ke...	6.956	0.000	9653251	0	1.122	N.D. #
23)	Chlordane-1	4.167	0.000	7365265	0	23.265	N.D. #
27)	Chlordane-5	6.077	6.797	17047224	3961138	52.561	42.821



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL090618\
Data File : PL039722.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Sep 2018 14:57
Operator : AJ\SJ
Sample : J4781-01
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 07 04:09:42 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL081318.M
Quant Title : GC Extractables
QLast Update : Wed Aug 15 15:04:22 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.302 min
Delta R.T.: -0.004 min
Response: 182083489
Conc: 26.05 ng/ml

Instrument :
ECD_L
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.905 min
Delta R.T.: -0.002 min
Response: 79647465
Conc: 29.71 ng/ml




#5 Aldrin

R.T.: 4.513 min
Delta R.T.: -0.016 min
Response: 6607702
Conc: 0.66 ng/ml



#5 Aldrin


R.T.: 0.000 min
Exp R.T.: 5.372 min
Response: 0
Conc: N.D.



#6 beta-BHC


R.T.: 4.219 min
Delta R.T.: -0.045 min
Response: 15856648
Conc: 3.41 ng/ml

Instrument :
ECD_L
ClientSampleId :




#6 beta-BHC

R.T.: 0.000 min
Exp R.T. : 4.726 min
Response: 0
Conc: N.D.




#8 Heptachlor epoxide

R.T.: 5.025 min
Delta R.T.: 0.059 min
Response: 20413355
Conc: 2.14 ng/ml



#8 Heptachlor epoxide


R.T.: 0.000 min
Exp R.T. : 5.751 min
Response: 0
Conc: N.D.



#14 Endrin


R.T.: 5.772 min
Delta R.T.: -0.015 min
Response: 8764928
Conc: 0.98 ng/ml

Instrument :
ECD_L
ClientSampleId :




#14 Endrin

R.T.: 0.000 min
Exp R.T. : 6.570 min
Response: 0
Conc: N.D.




#15 Endosulfan II

R.T.: 6.077 min
Delta R.T.: 0.016 min
Response: 17047224
Conc: 1.96 ng/ml



#15 Endosulfan II


R.T.: 6.797 min
Delta R.T.: 0.024 min
Response: 3961138
Conc: 1.54 ng/ml



#16 4,4'-DDD


R.T.: 5.929 min
Delta R.T.: 0.007 min
Response: 6995758
Conc: 0.88 ng/ml

Instrument :
ECD_L
ClientSampleId :




#16 4,4'-DDD

R.T.: 0.000 min
Exp R.T. : 6.693 min
Response: 0
Conc: N.D.




#20 Methoxychlor

R.T.: 6.704 min
Delta R.T.: 0.000 min
Response: 6922135
Conc: 1.69 ng/ml



#20 Methoxychlor


R.T.: 0.000 min
Exp R.T. : 7.446 min
Response: 0
Conc: N.D.



#21 Endrin ketone


R.T.: 6.956 min
Delta R.T.: 0.035 min
Response: 9653251
Conc: 1.12 ng/ml

Instrument :
ECD_L
ClientSampleId :




#21 Endrin ketone

R.T.: 0.000 min
Exp R.T. : 7.580 min
Response: 0
Conc: N.D.




#23 Chlordane-1

R.T.: 4.167 min
Delta R.T.: 0.023 min
Response: 7365265
Conc: 23.26 ng/ml



#23 Chlordane-1


R.T.: 0.000 min
Exp R.T. : 4.885 min
Response: 0
Conc: N.D.



#27 Chlordane-5


R.T.: 6.077 min
Delta R.T.: 0.013 min
Response: 17047224
Conc: 52.56 ng/ml

Instrument :
ECD_L
ClientSampleId :




#27 Chlordane-5

R.T.: 6.797 min
Delta R.T.: -0.052 min
Response: 3961138
Conc: 42.82 ng/ml



#28 Decachlorobiphenyl

R.T.: 7.955 min
Delta R.T.: -0.009 min
Response: 168518829
Conc: 23.99 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.891 min
Delta R.T.: -0.005 min
Response: 66495723
Conc: 28.65 ng/ml