

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL090619\
 Data File : PL052278.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Sep 2019 19:01
 Operator : SG\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Manual Integrations
 APPROVED

Ankita
 9/9/2019 3:07:21 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 06 23:03:17 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL081919.M
 Quant Title : GC Extractables
 QLast Update : Mon Aug 19 18:03:07 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.313	3.924	470.5E6	117.2E6	54.974	53.665
28)	SA Decachlor...	7.971	8.951	486.2E6	124.5E6	40.059m	40.900
Target Compounds							
2)	A alpha-BHC	3.740	4.312	787.0E6	172.6E6	54.314	53.555
3)	MA gamma-BHC...	4.016	4.595	744.0E6	166.9E6	56.550	53.019
4)	MA Heptachlor	4.299	5.096	808.3E6	180.0E6	53.391	51.332
5)	MB Aldrin	4.537	5.398	783.7E6	167.8E6	54.235	51.185
6)	B beta-BHC	4.271	4.773	311.3E6	78758784	54.526	54.040
7)	B delta-BHC	4.463	4.986	806.2E6	170.0E6	54.221	54.307
8)	B Heptachlo...	4.972	5.785	726.6E6	162.2E6	51.954	48.048
9)	A Endosulfan I	5.303	6.140	680.9E6	154.6E6	50.648	47.823
10)	B gamma-Chl...	5.195	6.018	769.3E6	170.0E6	51.520	46.575
11)	B alpha-Chl...	5.250	6.090	741.7E6	167.4E6	51.250	48.513
12)	B 4,4'-DDE	5.418	6.250	641.5E6	165.8E6	46.877	47.096
13)	MA Dieldrin	5.541	6.396	731.8E6	163.9E6	48.128	47.852
14)	MA Endrin	5.794	6.613	644.6E6	139.6E6	44.196	46.158
15)	B Endosulfa...	6.069	6.825	604.4E6	143.0E6	43.441	43.383
16)	A 4,4'-DDD	5.928	6.741	546.8E6	127.0E6	44.650	46.924
17)	MA 4,4'-DDT	6.162	7.038	544.2E6	127.5E6	42.680	43.539
18)	B Endrin al...	6.236	6.950	495.5E6	122.0E6	42.594	43.166
19)	B Endosulfa...	6.447	7.172	541.6E6	133.7E6	41.633	43.304
20)	A Methoxychlor	6.706	7.499	274.9E6	69471943	40.540	41.421
21)	B Endrin ke...	6.927	7.641	566.8E6	138.6E6	40.064m	43.315
22)	Mirex	7.101	8.088	398.7E6	101.0E6	37.662	41.521

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL090619\
Data File : PL052278.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Sep 2019 19:01
Operator : SG\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Manual Integrations
APPROVED

Ankita
9/9/2019 3:07:21 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 06 23:03:17 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL081919.M
Quant Title : GC Extractables
QLast Update : Mon Aug 19 18:03:07 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

