

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL090921\
 Data File : PL069773.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 09 Sep 2021 10:32
 Operator : AR\AJ
 Sample : M3263-02
 Misc : TOX MDL 100 PPB
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 MDL-WATER-QT4-2021-08

Manual Integrations
 APPROVED

mohammad
 9/10/2021 10:58:53 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 10 07:44:48 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL090721CLP.M
 Quant Title : GC Extractables
 QLast Update : Wed Sep 08 10:30:41 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml


System Monitoring Compounds							
1)	SA Tetrachlo...	4.633	5.569	7976357	23858733	18.764	17.534
27)	SA Decachlor...	10.402	11.496	26926644	41331620	35.402	35.276
Target Compounds							
22)	Toxaphene-1	7.731	8.620	399495	867967	121.116m	108.355m
23)	Toxaphene-2	7.903	9.154	462332	1169555	103.895m	85.812m
24)	Toxaphene-3	8.402	9.379	1164644	1083972	96.576m	101.171m
25)	Toxaphene-4	8.671	9.553	1173815	2044372	90.479m	97.498m
26)	Toxaphene-5	9.190	9.994	1472084	1455263	95.134	92.167m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL090921\
Data File : PL069773.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 Sep 2021 10:32
Operator : AR\AJ
Sample : M3263-02
Misc : TOX MDL 100 PPB
ALS Vial : 7 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 10 07:44:48 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL090721CLP.M
Quant Title : GC Extractables
QLast Update : Wed Sep 08 10:30:41 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm



Manual Integrations
APPROVED
mohammad
9/10/2021 10:58:53 AM

Instrument :
ECD_L
Client Sample :
MDL-WATER-QT4-2021-08