

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL091020\
 Data File : PL061367.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Sep 2020 17:30
 Operator : DD\AJ
 Sample : L3941-03
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 RBR-200031

Manual Integrations
 APPROVED

Ankita
 9/11/2020 11:43:08 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 10 18:37:01 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL090820.M
 Quant Title : GC Extractables
 QLast Update : Wed Sep 09 01:27:38 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.459	4.008	15817153	34772181	9.171m	11.304
28) SA Decachlor...	8.211	9.092	13496581	14319479	8.090	6.056 #

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL091020\
Data File : PL061367.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Sep 2020 17:30
Operator : DD\AJ
Sample : L3941-03
Misc :
ALS Vial : 18 Sample Multiplier: 1



Instrument :
ECD_L
ClientSampleId :
RBR-200031


Manual Integrations
APPROVED

Ankita
9/11/2020 11:43:08 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 10 18:37:01 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL090820.M
Quant Title : GC Extractables
QLast Update : Wed Sep 09 01:27:38 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene

R.T.: 3.459 min
Delta R.T.: 0.000 min
Response: 15817153
Conc: 9.17 ng/ml m

Instrument :
ECD_L
ClientSampleId :
RBR-200031


Manual Integrations
APPROVED

Ankita
9/11/2020 11:43:08 AM




#1 Tetrachloro-m-xylene

R.T.: 4.008 min
Delta R.T.: 0.000 min
Response: 34772181
Conc: 11.30 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.211 min
Delta R.T.: 0.001 min
Response: 13496581
Conc: 8.09 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.092 min
Delta R.T.: 0.000 min
Response: 14319479
Conc: 6.06 ng/ml