

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL091218\
 Data File : PL039862.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Sep 2018 19:15
 Operator : AJ\SJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Manual Integrations
 APPROVED

Sohil
 9/13/2018 11:55:27 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 13 02:40:11 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL081318.M
 Quant Title : GC Extractables
 QLast Update : Wed Aug 15 15:04:22 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.301	3.905	326.0E6	142.9E6	46.630	53.299
28)	SA Decachlor...	7.953	8.890	331.0E6	121.1E6	47.128	52.180
Target Compounds							
2)	A alpha-BHC	3.728	4.282	533.5E6	217.7E6	45.136	51.687
3)	MA gamma-BHC...	4.004	4.561	486.7E6	196.2E6	45.891	48.728
4)	MA Heptachlor	4.285	5.068	467.9E6	192.6E6	44.444	53.854
5)	MB Aldrin	4.522	5.368	467.5E6	181.4E6	46.671	51.229
6)	B beta-BHC	4.259	4.724	205.8E6	89476864	44.293	51.085
7)	B delta-BHC	4.450	4.936	494.9E6	198.3E6	45.596	49.804
8)	B Heptachlo...	4.956	5.747	439.0E6	170.9E6	45.975m	50.809
9)	A Endosulfan I	5.286	6.099	413.0E6	150.7E6	47.167	51.374
10)	B gamma-Chl...	5.179	5.979	457.6E6	169.7E6	46.327	52.431
11)	B alpha-Chl...	5.235	6.052	440.6E6	167.6E6	45.860	51.794
12)	B 4,4'-DDE	5.402	6.208	441.8E6	163.2E6	46.140m	52.776
13)	MA Dieldrin	5.524	6.352	457.2E6	156.7E6	46.882m	51.736
14)	MA Endrin	5.777	6.566	389.4E6	124.7E6	43.505	48.512
15)	B Endosulfa...	6.052	6.769	403.3E6	132.6E6	46.382	51.641
16)	A 4,4'-DDD	5.912	6.690	373.0E6	123.4E6	46.801m	55.121
17)	MA 4,4'-DDT	6.146	6.988	336.9E6	111.2E6	44.445	47.223
18)	B Endrin al...	6.218	6.892	312.3E6	112.3E6	46.076m	51.728
19)	B Endosulfa...	6.431	7.114	371.9E6	123.3E6	45.701	51.425
20)	A Methoxychlor	6.693	7.443	181.7E6	59272602	44.402	47.857
21)	B Endrin ke...	6.911	7.576	412.2E6	123.6E6	47.893m	50.015

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL091218\
Data File : PL039862.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Sep 2018 19:15
Operator : AJ\SJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Manual Integrations
APPROVED

Sohil
9/13/2018 11:55:27 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 13 02:40:11 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL081318.M
Quant Title : GC Extractables
QLast Update : Wed Aug 15 15:04:22 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

