

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL091219\
 Data File : PL052403.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Sep 2019 17:17
 Operator : SG\AJ
 Sample : K4852-16MS
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 BUS-2(0-5)MS

Manual Integrations
 APPROVED

Ankita
 9/13/2019 3:45:06 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 13 01:12:51 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL091119.M
 Quant Title : GC Extractables
 QLast Update : Thu Sep 12 03:37:35 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.313	3.923	189.3E6	49728904	21.012	22.835
28)	SA Decachlor...	7.969	8.947	249.1E6	54887673	17.754	18.499
Target Compounds							
2)	A alpha-BHC	3.740	4.311	897.4E6	184.1E6	59.329	57.226
3)	MA gamma-BHC...	4.015	4.594	831.3E6	183.5E6	58.298	58.680
4)	MA Heptachlor	4.297	5.094	945.3E6	204.2E6	59.491	58.257
5)	MB Aldrin	4.535	5.396	896.9E6	198.4E6	61.330	57.877
6)	B beta-BHC	4.270	4.771	359.1E6	87136136	60.590	55.392
7)	B delta-BHC	4.462	4.984	656.7E6	120.8E6	42.525	38.102
8)	B Heptachlo...	4.971	5.783	844.5E6	183.6E6	57.275	51.948
9)	A Endosulfan I	5.301	6.137	817.1E6	182.0E6	55.051	53.554
10)	B gamma-Chl...	5.193	6.016	900.7E6	197.3E6	54.977	51.958
11)	B alpha-Chl...	5.249	6.088	872.1E6	196.5E6	54.860	53.824
12)	B 4,4'-DDE	5.416	6.249	905.0E6	201.1E6	56.342	54.809
13)	MA Dieldrin	5.539	6.394	908.0E6	197.5E6	54.320	54.858
14)	MA Endrin	5.792	6.609	859.2E6	169.8E6	53.555	54.191m
15)	B Endosulfa...	6.067	6.824	782.9E6	177.5E6	50.309	51.549
16)	A 4,4'-DDD	5.927	6.739	720.8E6	160.8E6	53.314	55.044
17)	MA 4,4'-DDT	6.160	7.035	754.8E6	174.4E6	53.615	54.903
18)	B Endrin al...	6.234	6.947	638.3E6	150.6E6	50.062	50.982
19)	B Endosulfa...	6.445	7.170	734.5E6	168.2E6	49.579	51.531
20)	A Methoxychlor	6.704	7.496	392.4E6	90220352	53.819	52.792
21)	B Endrin ke...	6.926	7.638	845.5E6	187.3E6	51.644	54.714
22)	Mirex	7.099	8.085	593.5E6	134.1E6	49.234	49.871

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL091219\
Data File : PL052403.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Sep 2019 17:17
Operator : SG\AJ
Sample : K4852-16MS
Misc :
ALS Vial : 28 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampled :
BUS-2(0-5)MS

Manual Integrations
APPROVED

Ankita
9/13/2019 3:45:06 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 13 01:12:51 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL091119.M
Quant Title : GC Extractables
QLast Update : Thu Sep 12 03:37:35 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

