

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL091223\
 Data File : PL085254.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 Sep 2023 18:44
 Operator : AR\AJ
 Sample : TOXAPH501
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 TOXAPH5275

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 12 02:31:48 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL091223CLP.M
 Quant Title : GC Extractables
 QLast Update : Tue Sep 12 02:29:03 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.475	2.683	220.6E6	103.2E6	79.492	79.541
27)	SA Decachlor...	9.026	7.862	256.8E6	232.2E6	150.875	154.289
Target Compounds							
22)	Toxaphene-1	6.190	5.251	134.7E6	52560791	7989.760	7991.086
23)	Toxaphene-2	7.015	5.612	489.9E6	71962674	8090.766	7941.415
24)	Toxaphene-3	7.104	6.656	351.3E6	351.4E6	8087.011	8082.796
25)	Toxaphene-4	7.202	6.972	196.4E6	229.8E6	8052.238	8136.019
26)	Toxaphene-5	7.891	7.106	245.3E6	145.7E6	8090.969	8335.487

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL091223\
Data File : PL085254.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Sep 2023 18:44
Operator : AR\AJ
Sample : TOXAPH501
Misc :
ALS Vial : 8 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 12 02:31:48 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL091223CLP.M
Quant Title : GC Extractables
QLast Update : Tue Sep 12 02:29:03 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

