

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL091420\
 Data File : PL061411.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Sep 2020 12:45
 Operator : DD\AJ
 Sample : L3869-01
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 OR-03-091120

Manual Integrations
 APPROVED

Ankita
 9/15/2020 10:17:21 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 14 16:32:32 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL090820.M
 Quant Title : GC Extractables
 QLast Update : Wed Sep 09 01:27:38 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.458	4.007	27248789	44668660	15.799	14.522
28)	SA Decachlor...	8.210	9.093	21999303	34918606	13.186	14.767
Target Compounds							
10)	B gamma-Chl...	5.396	6.119	3603394	5245220	1.945m	1.521m
11)	B alpha-Chl...	5.453	6.194	4484545	10934685	2.421m	3.142 #
12)	B 4,4'-DDE	5.617	6.346	1405238	2689400	0.826m	0.839m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL091420\
Data File : PL061411.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Sep 2020 12:45
Operator : DD\AJ
Sample : L3869-01
Misc :
ALS Vial : 8 Sample Multiplier: 1



Instrument :
ECD_L
ClientSampleId :
OR-03-091120


Manual Integrations
APPROVED

Ankita
9/15/2020 10:17:21 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 14 16:32:32 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL090820.M
Quant Title : GC Extractables
QLast Update : Wed Sep 09 01:27:38 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene

R.T.: 3.458 min
Delta R.T.: 0.000 min
Response: 27248789
Conc: 15.80 ng/ml

Instrument :
ECD_L
ClientSampleId :
OR-03-091120


Manual Integrations
APPROVED

Ankita
9/15/2020 10:17:21 AM




#1 Tetrachloro-m-xylene

R.T.: 4.007 min
Delta R.T.: -0.001 min
Response: 44668660
Conc: 14.52 ng/ml




#10 gamma-Chlordane

R.T.: 5.396 min
Delta R.T.: -0.001 min
Response: 3603394
Conc: 1.95 ng/ml m



#10 gamma-Chlordane

R.T.: 6.119 min
Delta R.T.: -0.003 min
Response: 5245220
Conc: 1.52 ng/ml m




#11 alpha-Chlordane

R.T.: 5.453 min
Delta R.T.: -0.002 min
Response: 4484545
Conc: 2.42 ng/ml m

Instrument :
ECD_L
ClientSampleId :
OR-03-091120


Manual Integrations
APPROVED

Ankita
9/15/2020 10:17:21 AM




#11 alpha-Chlordane

R.T.: 6.194 min
Delta R.T.: 0.000 min
Response: 10934685
Conc: 3.14 ng/ml




#12 4,4'-DDE

R.T.: 5.617 min
Delta R.T.: -0.004 min
Response: 1405238
Conc: 0.83 ng/ml m



#12 4,4'-DDE

R.T.: 6.346 min
Delta R.T.: -0.006 min
Response: 2689400
Conc: 0.84 ng/ml m




#28 Decachlorobiphenyl

R.T.: 8.210 min
Delta R.T.: 0.000 min
Response: 21999303
Conc: 13.19 ng/ml

Instrument :
ECD_L
ClientSampleId :
OR-03-091120

Manual Integrations
APPROVED

Ankita
9/15/2020 10:17:21 AM



#28 Decachlorobiphenyl

R.T.: 9.093 min
Delta R.T.: 0.000 min
Response: 34918606
Conc: 14.77 ng/ml