

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL091422\
 Data File : PL077753.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Sep 2022 17:05
 Operator : AR\AJ
 Sample : N4613-02MS
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 SASP2-T-6-(18-23)MS

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 14 18:10:41 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL091222.M
 Quant Title : GC Extractables
 QLast Update : Tue Sep 13 13:26:11 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.417	2.627	29533424	54104192	18.424	18.225
28)	SA Decachlor...	8.909	7.726	22803933	39351584	17.644	15.676
Target Compounds							
2)	A alpha-BHC	3.882	3.119	112.8E6	216.8E6	47.521	47.814
3)	MA gamma-BHC...	4.217	3.443	105.6E6	196.7E6	47.725	48.699
4)	MA Heptachlor	4.794	3.775	104.3E6	187.0E6	47.116	47.209
5)	MB Aldrin	5.133	4.049	100.1E6	192.2E6	50.354	50.649
6)	B beta-BHC	4.430	3.743	45278000	77692890	46.588	46.668
7)	B delta-BHC	4.674	3.966	85459706	170.5E6	41.623	42.015
8)	B Heptachlo...	5.567	4.549	89169776	167.0E6	51.150	49.392
9)	A Endosulfan I	5.953	4.914	79191086	152.5E6	49.135	48.335
10)	B gamma-Chl...	5.823	4.798	86597620	167.1E6	49.459	49.040
11)	B alpha-Chl...	5.899	4.860	84929788	161.9E6	49.168	48.878
12)	B 4,4'-DDE	6.079	5.054	80258558	160.2E6	50.418	49.245
13)	MA Dieldrin	6.231	5.177	81410816	169.2E6	50.439	51.028
14)	MA Endrin	6.463	5.450	65125002	143.8E6	47.090	50.437
15)	B Endosulfa...	6.691	5.746	68683112	139.5E6	49.721	52.937
16)	A 4,4'-DDD	6.604	5.607	59838906	132.0E6	51.107	50.955
17)	MA 4,4'-DDT	6.915	5.855	55439207	117.9E6	43.752	44.729
18)	B Endrin al...	6.823	5.925	55680222	108.6E6	52.654	52.616
19)	B Endosulfa...	7.055	6.148	61684753	127.4E6	48.169	47.384
20)	A Methoxychlor	7.396	6.434	30868124	62606726	46.738	45.717
21)	B Endrin ke...	7.544	6.647	68681451	146.7E6	49.704	48.442
22)	Mirex	8.004	6.824	52882391	104.0E6	50.365	48.355

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL091422\
Data File : PL077753.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Sep 2022 17:05
Operator : AR\AJ
Sample : N4613-02MS
Misc :
ALS Vial : 28 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
SASP2-T-6-(18-23)MS

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 14 18:10:41 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL091222.M
Quant Title : GC Extractables
QLast Update : Tue Sep 13 13:26:11 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

