

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL091423\
 Data File : PL085308.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 Sep 2023 13:44
 Operator : AR\AJ
 Sample : PB155491BS
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PB155491BS

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 13 22:35:29 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL090523.M
 Quant Title : GC Extractables
 QLast Update : Tue Sep 05 15:22:36 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.476	2.682	42547789	24467871	17.395	21.535
28)	SA Decachlor...	9.022	7.861	34822623	25559357	21.654	19.202
Target Compounds							
23)	Chlordane-1	4.643	3.683	19290854	6901899	193.864	177.605
24)	Chlordane-2	5.176	4.264	20520349	9252098	195.879	191.966
25)	Chlordane-3	5.890	4.897	70085018	24108684	193.903	179.543
26)	Chlordane-4	5.972	4.959	85298533	25580920	196.610	172.022
27)	Chlordane-5	6.822	5.857	14068702	9428054	201.505	185.209



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL091423\
Data File : PL085308.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Sep 2023 13:44
Operator : AR\AJ
Sample : PB155491BS
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PB155491BS

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 13 22:35:29 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL090523.M
Quant Title : GC Extractables
QLast Update : Tue Sep 05 15:22:36 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene


R.T.: 3.476 min
Delta R.T.: 0.000 min
Response: 42547789
Conc: 17.40 ng/ml

Instrument :
ECD_L
ClientSampleId :
PB155491BS




#1 Tetrachloro-m-xylene

R.T.: 2.682 min
Delta R.T.: -0.002 min
Response: 24467871
Conc: 21.53 ng/ml




#23 Chlordane-1

R.T.: 4.643 min
Delta R.T.: 0.000 min
Response: 19290854
Conc: 193.86 ng/ml



#23 Chlordane-1


R.T.: 3.683 min
Delta R.T.: -0.002 min
Response: 6901899
Conc: 177.61 ng/ml



#24 Chlordane-2


R.T.: 5.176 min
Delta R.T.: -0.001 min
Response: 20520349
Conc: 195.88 ng/ml

Instrument :
ECD_L
ClientSampleId :
PB155491BS




#24 Chlordane-2

R.T.: 4.264 min
Delta R.T.: -0.001 min
Response: 9252098
Conc: 191.97 ng/ml




#25 Chlordane-3

R.T.: 5.890 min
Delta R.T.: -0.002 min
Response: 70085018
Conc: 193.90 ng/ml



#25 Chlordane-3


R.T.: 4.897 min
Delta R.T.: -0.002 min
Response: 24108684
Conc: 179.54 ng/ml



#26 Chlordane-4


R.T.: 5.972 min
Delta R.T.: -0.002 min
Response: 85298533
Conc: 196.61 ng/ml

Instrument :
ECD_L
ClientSampleId :
PB155491BS




#26 Chlordane-4

R.T.: 4.959 min
Delta R.T.: -0.003 min
Response: 25580920
Conc: 172.02 ng/ml




#27 Chlordane-5

R.T.: 6.822 min
Delta R.T.: -0.003 min
Response: 14068702
Conc: 201.50 ng/ml



#27 Chlordane-5


R.T.: 5.857 min
Delta R.T.: -0.003 min
Response: 9428054
Conc: 185.21 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.022 min
Delta R.T.: -0.005 min
Response: 34822623
Conc: 21.65 ng/ml

Instrument :
ECD_L
ClientSampleId :
PB155491BS



#28 Decachlorobiphenyl

R.T.: 7.861 min
Delta R.T.: -0.004 min
Response: 25559357
Conc: 19.20 ng/ml