

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL091518\
 Data File : PL039970.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Sep 2018 13:37
 Operator : AJ\SJ
 Sample : J4910-05
 Misc :
 ALS Vial : 77 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 BZTRANS-5

Manual Integrations
 APPROVED

Sohil
 9/17/2018 11:15:27 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 17 00:54:35 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL091418.M
 Quant Title : GC Extractables
 QLast Update : Fri Sep 14 23:10:39 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.300	3.904	115.9E6	48095168	14.949	15.073
28)	SA Decachlor...	7.951	8.886	145.3E6	51206647	17.437	18.853m

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL091518\
Data File : PL039970.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Sep 2018 13:37
Operator : AJ\SJ
Sample : J4910-05
Misc :
ALS Vial : 77 Sample Multiplier: 1



Instrument :
ECD_L
ClientSampleId :
BZTRANS-5


Manual Integrations
APPROVED

Sohil
9/17/2018 11:15:27 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 17 00:54:35 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL091418.M
Quant Title : GC Extractables
QLast Update : Fri Sep 14 23:10:39 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene

R.T.: 3.300 min
Delta R.T.: 0.000 min
Response: 115928724
Conc: 14.95 ng/ml

Instrument :
ECD_L
ClientSampleId :
BZTRANS-5


Manual Integrations
APPROVED

Sohil
9/17/2018 11:15:27 AM




#1 Tetrachloro-m-xylene

R.T.: 3.904 min
Delta R.T.: 0.000 min
Response: 48095168
Conc: 15.07 ng/ml



#28 Decachlorobiphenyl

R.T.: 7.951 min
Delta R.T.: 0.000 min
Response: 145296813
Conc: 17.44 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.886 min
Delta R.T.: -0.002 min
Response: 51206647
Conc: 18.85 ng/ml m