

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL091624\
 Data File : PL091775.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 16 Sep 2024 16:53
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 17 01:26:31 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL090924.M
 Quant Title : GC Extractables
 QLast Update : Mon Sep 09 19:26:36 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.543	2.782	87400002	104.9E6	56.076	56.087
28)	SA Decachlor...	9.060	7.923	53378684	88983318	43.921	48.189
Target Compounds							
2)	A alpha-BHC	3.999	3.285	118.8E6	150.9E6	55.008	56.754
3)	MA gamma-BHC...	4.331	3.615	111.0E6	140.3E6	53.973	57.048
4)	MA Heptachlor	4.921	3.954	105.3E6	134.9E6	56.234	56.026
5)	MB Aldrin	5.262	4.235	96456019	129.3E6	51.786	56.657
6)	B beta-BHC	4.528	3.914	50691724	60505356	55.704	55.384
7)	B delta-BHC	4.776	4.143	107.7E6	134.8E6	56.078	58.599
8)	B Heptachlo...	5.688	4.737	89556130	113.1E6	53.668	53.682
9)	A Endosulfan I	6.074	5.107	82249631	109.5E6	52.788	56.754
10)	B gamma-Chl...	5.944	4.988	88712623	120.4E6	53.097	54.572
11)	B alpha-Chl...	6.024	5.051	88807565	117.8E6	55.097	56.925
12)	B 4,4'-DDE	6.197	5.240	79603389	110.2E6	54.573	58.080
13)	MA Dieldrin	6.349	5.372	86613235	112.8E6	53.080	54.890
14)	MA Endrin	6.578	5.647	61590323	92587012	46.319	49.061
15)	B Endosulfa...	6.798	5.942	77764812	101.1E6	50.171	57.172
16)	A 4,4'-DDD	6.714	5.795	70041546	91569611	55.258	54.644
17)	MA 4,4'-DDT	7.027	6.046	62115618	84491005	47.353	52.584
18)	B Endrin al...	6.928	6.121	62612543	77209087	52.698	56.479
19)	B Endosulfa...	7.162	6.344	70354894	94375252	50.383	54.994
20)	A Methoxychlor	7.503	6.621	34198618	43507495	50.555	49.119
21)	B Endrin ke...	7.647	6.849	79172516	111.1E6	51.201	55.240
22)	Mirex	8.121	7.031	58621226	89036691	48.956	50.211
24)	Chlordane-2	5.262f	4.372	96456019	4286608	1254.595	57.192 #
25)	Chlordane-3	5.944	4.988	88712623	120.4E6	345.321	516.290 #
26)	Chlordane-4	6.024	5.051	88807565	117.8E6	299.867	564.791 #
27)	Chlordane-5	0.000	5.942	0	101.1E6	N.D.	1411.279 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL091624\
Data File : PL091775.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 16 Sep 2024 16:53
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 17 01:26:31 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL090924.M
Quant Title : GC Extractables
QLast Update : Mon Sep 09 19:26:36 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene


R.T.: 3.543 min
Delta R.T.: 0.000 min
Response: 87400002
Conc: 56.08 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#1 Tetrachloro-m-xylene

R.T.: 2.782 min
Delta R.T.: 0.000 min
Response: 104899131
Conc: 56.09 ng/ml




#2 alpha-BHC

R.T.: 3.999 min
Delta R.T.: 0.000 min
Response: 118814021
Conc: 55.01 ng/ml



#2 alpha-BHC


R.T.: 3.285 min
Delta R.T.: -0.001 min
Response: 150940145
Conc: 56.75 ng/ml



#3 gamma-BHC (Lindane)


R.T.: 4.331 min
Delta R.T.: 0.000 min
Response: 110952918
Conc: 53.97 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#3 gamma-BHC (Lindane)

R.T.: 3.615 min
Delta R.T.: 0.000 min
Response: 140306705
Conc: 57.05 ng/ml




#4 Heptachlor

R.T.: 4.921 min
Delta R.T.: 0.000 min
Response: 105328506
Conc: 56.23 ng/ml



#4 Heptachlor


R.T.: 3.954 min
Delta R.T.: -0.002 min
Response: 134927649
Conc: 56.03 ng/ml



#5 Aldrin


R.T.: 5.262 min
Delta R.T.: 0.000 min
Response: 96456019
Conc: 51.79 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#5 Aldrin

R.T.: 4.235 min
Delta R.T.: -0.001 min
Response: 129264248
Conc: 56.66 ng/ml




#6 beta-BHC

R.T.: 4.528 min
Delta R.T.: 0.000 min
Response: 50691724
Conc: 55.70 ng/ml



#6 beta-BHC


R.T.: 3.914 min
Delta R.T.: -0.001 min
Response: 60505356
Conc: 55.38 ng/ml



#7 delta-BHC


R.T.: 4.776 min
Delta R.T.: 0.000 min
Response: 107735622
Conc: 56.08 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#7 delta-BHC

R.T.: 4.143 min
Delta R.T.: -0.002 min
Response: 134821379
Conc: 58.60 ng/ml




#8 Heptachlor epoxide

R.T.: 5.688 min
Delta R.T.: 0.000 min
Response: 89556130
Conc: 53.67 ng/ml



#8 Heptachlor epoxide


R.T.: 4.737 min
Delta R.T.: -0.002 min
Response: 113147303
Conc: 53.68 ng/ml



#9 Endosulfan I


R.T.: 6.074 min
Delta R.T.: 0.000 min
Response: 82249631
Conc: 52.79 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#9 Endosulfan I

R.T.: 5.107 min
Delta R.T.: -0.002 min
Response: 109497995
Conc: 56.75 ng/ml




#10 gamma-Chlordane

R.T.: 5.944 min
Delta R.T.: 0.000 min
Response: 88712623
Conc: 53.10 ng/ml



#10 gamma-Chlordane


R.T.: 4.988 min
Delta R.T.: -0.002 min
Response: 120368399
Conc: 54.57 ng/ml



#11 alpha-Chlordane


R.T.: 6.024 min
Delta R.T.: 0.000 min
Response: 88807565
Conc: 55.10 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#11 alpha-Chlordane

R.T.: 5.051 min
Delta R.T.: -0.002 min
Response: 117777522
Conc: 56.92 ng/ml




#12 4,4'-DDE

R.T.: 6.197 min
Delta R.T.: 0.000 min
Response: 79603389
Conc: 54.57 ng/ml



#12 4,4'-DDE


R.T.: 5.240 min
Delta R.T.: -0.002 min
Response: 110161618
Conc: 58.08 ng/ml



#13 Dieldrin


R.T.: 6.349 min
Delta R.T.: 0.000 min
Response: 86613235
Conc: 53.08 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#13 Dieldrin

R.T.: 5.372 min
Delta R.T.: -0.002 min
Response: 112823727
Conc: 54.89 ng/ml




#14 Endrin

R.T.: 6.578 min
Delta R.T.: 0.000 min
Response: 61590323
Conc: 46.32 ng/ml



#14 Endrin


R.T.: 5.647 min
Delta R.T.: -0.002 min
Response: 92587012
Conc: 49.06 ng/ml



#15 Endosulfan II


R.T.: 6.798 min
Delta R.T.: 0.000 min
Response: 77764812
Conc: 50.17 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#15 Endosulfan II

R.T.: 5.942 min
Delta R.T.: -0.001 min
Response: 101133077
Conc: 57.17 ng/ml




#16 4,4'-DDD

R.T.: 6.714 min
Delta R.T.: 0.000 min
Response: 70041546
Conc: 55.26 ng/ml



#16 4,4'-DDD


R.T.: 5.795 min
Delta R.T.: -0.002 min
Response: 91569611
Conc: 54.64 ng/ml



#17 4,4'-DDT


R.T.: 7.027 min
Delta R.T.: 0.000 min
Response: 62115618
Conc: 47.35 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#17 4,4'-DDT

R.T.: 6.046 min
Delta R.T.: -0.002 min
Response: 84491005
Conc: 52.58 ng/ml




#18 Endrin aldehyde

R.T.: 6.928 min
Delta R.T.: 0.000 min
Response: 62612543
Conc: 52.70 ng/ml



#18 Endrin aldehyde


R.T.: 6.121 min
Delta R.T.: -0.002 min
Response: 77209087
Conc: 56.48 ng/ml



#19 Endosulfan Sulfate


R.T.: 7.162 min
Delta R.T.: 0.000 min
Response: 70354894
Conc: 50.38 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#19 Endosulfan Sulfate

R.T.: 6.344 min
Delta R.T.: -0.002 min
Response: 94375252
Conc: 54.99 ng/ml




#20 Methoxychlor

R.T.: 7.503 min
Delta R.T.: 0.000 min
Response: 34198618
Conc: 50.55 ng/ml



#20 Methoxychlor


R.T.: 6.621 min
Delta R.T.: -0.002 min
Response: 43507495
Conc: 49.12 ng/ml



#21 Endrin ketone


R.T.: 7.647 min
Delta R.T.: 0.000 min
Response: 79172516
Conc: 51.20 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#21 Endrin ketone

R.T.: 6.849 min
Delta R.T.: -0.002 min
Response: 111109612
Conc: 55.24 ng/ml




#22 Mirex

R.T.: 8.121 min
Delta R.T.: 0.000 min
Response: 58621226
Conc: 48.96 ng/ml



#22 Mirex


R.T.: 7.031 min
Delta R.T.: -0.002 min
Response: 89036691
Conc: 50.21 ng/ml



#24 Chlordane-2


R.T.: 5.262 min
Delta R.T.: 0.027 min
Response: 96456019
Conc: 1254.60 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#24 Chlordane-2

R.T.: 4.372 min
Delta R.T.: 0.013 min
Response: 4286608
Conc: 57.19 ng/ml




#25 Chlordane-3

R.T.: 5.944 min
Delta R.T.: -0.001 min
Response: 88712623
Conc: 345.32 ng/ml



#25 Chlordane-3


R.T.: 4.988 min
Delta R.T.: -0.002 min
Response: 120368399
Conc: 516.29 ng/ml



#26 Chlordane-4


R.T.: 6.024 min
Delta R.T.: -0.003 min
Response: 88807565
Conc: 299.87 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#26 Chlordane-4

R.T.: 5.051 min
Delta R.T.: -0.001 min
Response: 117777522
Conc: 564.79 ng/ml




#27 Chlordane-5

R.T.: 0.000 min
Exp R.T. : 6.876 min
Response: 0
Conc: N.D.



#27 Chlordane-5


R.T.: 5.942 min
Delta R.T.: -0.007 min
Response: 101133077
Conc: 1411.28 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.060 min
Delta R.T.: 0.000 min
Response: 53378684
Conc: 43.92 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050



#28 Decachlorobiphenyl

R.T.: 7.923 min
Delta R.T.: -0.002 min
Response: 88983318
Conc: 48.19 ng/ml