

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL091819\
 Data File : PL052562.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Sep 2019 11:52
 Operator : SG\AJ
 Sample : PB123112BL
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 19 01:48:36 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL091719.M
 Quant Title : GC Extractables
 QLast Update : Tue Sep 17 11:25:25 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.557	4.046	167.9E6	48747764	21.537	22.372
28)	SA Decachlor...	8.358	9.123	173.3E6	56432080	19.823	23.566
Target Compounds							
7)	B delta-BHC	0.000	5.086f	0	722578	N.D.	0.252 #
8)	B Heptachlo...	0.000	5.918	0	2085050	N.D.	0.795 #


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL091819\
Data File : PL052562.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Sep 2019 11:52
Operator : SG\AJ
Sample : PB123112BL
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 19 01:48:36 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL091719.M
Quant Title : GC Extractables
QLast Update : Tue Sep 17 11:25:25 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene


R.T.: 3.557 min
Delta R.T.: 0.006 min
Response: 167937923
Conc: 21.54 ng/ml

Instrument :
ECD_L
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 4.046 min
Delta R.T.: 0.000 min
Response: 48747764
Conc: 22.37 ng/ml




#7 delta-BHC

R.T.: 0.000 min
Exp R.T. : 4.743 min
Response: 0
Conc: N.D.



#7 delta-BHC


R.T.: 5.086 min
Delta R.T.: -0.021 min
Response: 722578
Conc: 0.25 ng/ml



#8 Heptachlor epoxide


R.T.: 0.000 min
Exp R.T.: 5.293 min
Response: 0
Conc: N.D.

Instrument :
ECD_L
ClientSampleId :




#8 Heptachlor epoxide

R.T.: 5.918 min
Delta R.T.: -0.004 min
Response: 2085050
Conc: 0.80 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.358 min
Delta R.T.: 0.004 min
Response: 173341819
Conc: 19.82 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.123 min
Delta R.T.: 0.004 min
Response: 56432080
Conc: 23.57 ng/ml