

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL091819\
 Data File : PL052584.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Sep 2019 17:29
 Operator : SG\AJ
 Sample : PB123206TB
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 19 01:52:33 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL091719.M
 Quant Title : GC Extractables
 QLast Update : Tue Sep 17 11:25:25 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.551	4.046	162.8E6	49202117	20.873	22.581
28)	SA Decachlor...	8.350	9.119	167.7E6	56178947	19.180	23.460
Target Compounds							
7)	B delta-BHC	0.000	5.083f	0	982045	N.D.	0.343 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL091819\
Data File : PL052584.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Sep 2019 17:29
Operator : SG\AJ
Sample : PB123206TB
Misc :
ALS Vial : 28 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 19 01:52:33 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL091719.M
Quant Title : GC Extractables
QLast Update : Tue Sep 17 11:25:25 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene


R.T.: 3.551 min
Delta R.T.: 0.000 min
Response: 162764851
Conc: 20.87 ng/ml

Instrument :
ECD_L
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 4.046 min
Delta R.T.: 0.000 min
Response: 49202117
Conc: 22.58 ng/ml




#7 delta-BHC

R.T.: 0.000 min
Exp R.T. : 4.743 min
Response: 0
Conc: N.D.



#7 delta-BHC

R.T.: 5.083 min
Delta R.T.: -0.024 min
Response: 982045
Conc: 0.34 ng/ml



#28 Decachlorobiphenyl


R.T.: 8.350 min

Delta R.T.: -0.004 min

Response: 167718354

Conc: 19.18 ng/ml

Instrument :
ECD_L
ClientSampleId :



#28 Decachlorobiphenyl

R.T.: 9.119 min

Delta R.T.: 0.000 min

Response: 56178947

Conc: 23.46 ng/ml