

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL091918\
 Data File : PL040117.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 Sep 2018 16:12
 Operator : AJ\SJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Manual Integrations
 APPROVED

Sohil
 9/20/2018 2:02:53 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 20 01:15:03 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL091418.M
 Quant Title : GC Extractables
 QLast Update : Fri Sep 14 23:10:39 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.300	3.903	386.2E6	159.7E6	49.807	50.051
28)	SA Decachlor...	7.952	8.891	395.1E6	139.0E6	47.414	51.159
Target Compounds							
2)	A alpha-BHC	3.727	4.281	639.3E6	251.6E6	50.395	48.617
3)	MA gamma-BHC...	4.002	4.559	583.7E6	227.0E6	51.000	49.856
4)	MA Heptachlor	4.284	5.066	569.5E6	224.0E6	49.364	51.941
5)	MB Aldrin	4.520	5.367	565.1E6	215.1E6	51.268	49.525
6)	B beta-BHC	4.258	4.722	253.7E6	102.5E6	51.579	52.250
7)	B delta-BHC	4.450	4.935	606.9E6	222.8E6	51.609	56.009
8)	B Heptachlo...	4.957	5.746	525.9E6	202.6E6	51.607	51.227
9)	A Endosulfan I	5.286	6.097	484.2E6	174.5E6	48.867	49.138
10)	B gamma-Chl...	5.178	5.978	553.8E6	199.7E6	49.705	52.754
11)	B alpha-Chl...	5.233	6.051	528.8E6	192.9E6	49.500	48.698
12)	B 4,4'-DDE	5.402	6.208	533.6E6	185.7E6	50.347	48.822
13)	MA Dieldrin	5.524	6.351	554.4E6	182.5E6	49.990	47.974
14)	MA Endrin	5.777	6.565	471.3E6	145.9E6	45.925	44.821
15)	B Endosulfa...	6.052	6.768	503.3E6	150.9E6	52.535	45.681
16)	A 4,4'-DDD	5.912	6.689	455.2E6	142.3E6	52.501	49.207
17)	MA 4,4'-DDT	6.145	6.987	387.8E6	122.5E6	42.843	40.938
18)	B Endrin al...	6.219	6.891	377.7E6	128.4E6	49.271	48.890
19)	B Endosulfa...	6.430	7.113	447.9E6	139.9E6	48.875	47.637
20)	A Methoxychlor	6.693	7.443	208.0E6	64953458	44.880	44.299
21)	B Endrin ke...	6.910	7.576	525.1E6	143.9E6	53.772m	50.483

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL091918\
Data File : PL040117.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 Sep 2018 16:12
Operator : AJ\SJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Manual Integrations
APPROVED

Sohil
9/20/2018 2:02:53 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 20 01:15:03 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL091418.M
Quant Title : GC Extractables
QLast Update : Fri Sep 14 23:10:39 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

