

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL091919\
 Data File : PL052614.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 Sep 2019 09:34
 Operator : SG\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Manual Integrations
 APPROVED

Ankita
 9/20/2019 3:51:21 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 20 00:52:39 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL091719.M
 Quant Title : GC Extractables
 QLast Update : Tue Sep 17 11:25:25 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.551	4.046	410.3E6	118.4E6	52.617	54.351
28)	SA Decachlor...	8.350	9.118	375.5E6	123.2E6	42.945	51.444
Target Compounds							
2)	A alpha-BHC	3.994	4.437	663.0E6	171.6E6	55.147m	56.126
3)	MA gamma-BHC...	4.283	4.722	609.5E6	160.9E6	53.870	55.733
4)	MA Heptachlor	4.592	5.232	610.3E6	165.2E6	54.010	56.285
5)	MB Aldrin	4.845	5.537	572.9E6	149.0E6	53.050	56.245
6)	B beta-BHC	4.532	4.889	253.4E6	74972483	51.789	56.216
7)	B delta-BHC	4.741	5.107	606.0E6	158.3E6	53.627	55.312
8)	B Heptachlo...	5.292	5.922	526.5E6	154.1E6	51.533	58.752
9)	A Endosulfan I	5.634	6.279	491.9E6	128.8E6	50.420	55.259
10)	B gamma-Chl...	5.521	6.155	546.0E6	143.8E6	51.413	56.528
11)	B alpha-Chl...	5.580	6.228	528.3E6	139.1E6	51.102	55.667
12)	B 4,4'-DDE	5.743	6.384	491.4E6	134.5E6	52.666	55.151
13)	MA Dieldrin	5.879	6.535	527.5E6	134.5E6	50.884	55.062
14)	MA Endrin	6.138	6.753	446.2E6	104.0E6	53.971	53.729
15)	B Endosulfa...	6.412	6.961	457.9E6	119.6E6	47.211	53.568
16)	A 4,4'-DDD	6.262	6.872	395.0E6	103.8E6	51.574	56.357
17)	MA 4,4'-DDT	6.505	7.171	402.6E6	102.5E6	51.879	56.051
18)	B Endrin al...	6.580	7.084	368.3E6	99843499	47.633	53.823
19)	B Endosulfa...	6.793	7.307	383.4E6	108.6E6	46.221	52.887
20)	A Methoxychlor	7.052	7.628	200.3E6	55344110	49.546	54.617
21)	B Endrin ke...	7.287	7.778	465.5E6	115.9E6	44.829	52.208
22)	Mirex	7.484	8.233	343.0E6	90701830	44.849	52.997

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL091919\
Data File : PL052614.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 Sep 2019 09:34
Operator : SG\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampled :
PSTDCCC050

Manual Integrations
APPROVED

Ankita
9/20/2019 3:51:21 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 20 00:52:39 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL091719.M
Quant Title : GC Extractables
QLast Update : Tue Sep 17 11:25:25 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

