

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092021\
 Data File : PL069905.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Sep 2021 14:26
 Operator : AR\AJ
 Sample : M3786-01
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 CHRT-26878

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 21 02:32:40 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL091021.M
 Quant Title : GC Extractables
 QLast Update : Mon Sep 13 17:55:23 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.635	5.567	9478309	25019751	23.264	20.555
28)	SA Decachlor...	10.400	11.492	14648755	22616168	19.892	19.793
Target Compounds							
11)	B alpha-Chl...	7.417	8.385	333455	711480	0.513m	0.523
12)	B 4,4'-DDE	7.635	8.567	3615361	7386049	6.377	5.590m
16)	A 4,4'-DDD	8.220	9.109	4582501	10186668	9.285	9.569
17)	MA 4,4'-DDT	8.480	9.426	2542305	3928070	4.434	3.509



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092021\
Data File : PL069905.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Sep 2021 14:26
Operator : AR\AJ
Sample : M3786-01
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampled :
CHRT-26878

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 21 02:32:40 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL091021.M
Quant Title : GC Extractables
QLast Update : Mon Sep 13 17:55:23 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene


R.T.: 4.635 min
Delta R.T.: 0.002 min
Response: 9478309
Conc: 23.26 ng/ml

Instrument :
ECD_L
ClientSampleId :
CHRT-26878




#1 Tetrachloro-m-xylene

R.T.: 5.567 min
Delta R.T.: 0.000 min
Response: 25019751
Conc: 20.55 ng/ml




#11 alpha-Chlordane

R.T.: 7.417 min
Delta R.T.: -0.003 min
Response: 333455
Conc: 0.51 ng/ml m



#11 alpha-Chlordane


R.T.: 8.385 min
Delta R.T.: 0.004 min
Response: 711480
Conc: 0.52 ng/ml



#12 4,4'-DDE


R.T.: 7.635 min
Delta R.T.: 0.002 min
Response: 3615361
Conc: 6.38 ng/ml

Instrument :
ECD_L
ClientSampleId :
CHRT-26878




#12 4,4'-DDE

R.T.: 8.567 min
Delta R.T.: -0.001 min
Response: 7386049
Conc: 5.59 ng/ml m




#16 4,4'-DDD

R.T.: 8.220 min
Delta R.T.: 0.002 min
Response: 4582501
Conc: 9.29 ng/ml



#16 4,4'-DDD


R.T.: 9.109 min
Delta R.T.: 0.000 min
Response: 10186668
Conc: 9.57 ng/ml



#17 4,4'-DDT


R.T.: 8.480 min
Delta R.T.: 0.001 min
Response: 2542305
Conc: 4.43 ng/ml

Instrument :
ECD_L
ClientSampleId :
CHRT-26878




#17 4,4'-DDT

R.T.: 9.426 min
Delta R.T.: 0.000 min
Response: 3928070
Conc: 3.51 ng/ml



#28 Decachlorobiphenyl

R.T.: 10.400 min
Delta R.T.: 0.001 min
Response: 14648755
Conc: 19.89 ng/ml



#28 Decachlorobiphenyl

R.T.: 11.492 min
Delta R.T.: 0.000 min
Response: 22616168
Conc: 19.79 ng/ml