

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092022\
 Data File : PL077869.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Sep 2022 16:33
 Operator : AR\AJ
 Sample : N4703-02MS
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 ETGI-363MS

Manual Integrations
 APPROVED

Reviewed By :Abdul Mirza 09/21/2022
 Supervised By :Ankita Jodhani 09/21/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 21 02:53:23 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL091222.M
 Quant Title : GC Extractables
 QLast Update : Tue Sep 13 13:26:11 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.415	2.623	33268315	64789966	20.754	21.824
28)	SA Decachlor...	8.902	7.718	27013465	56002049	20.901m	22.310m
Target Compounds							
2)	A alpha-BHC	3.880	3.115	115.7E6	233.0E6	48.774	51.385
3)	MA gamma-BHC...	4.214	3.439	106.4E6	204.3E6	48.083m	50.600
4)	MA Heptachlor	4.791	3.771	104.2E6	203.4E6	47.062	51.351
5)	MB Aldrin	5.131	4.044	99975379	203.6E6	50.304	53.640
6)	B beta-BHC	4.429	3.739	48150135	82061757	49.543	49.292
7)	B delta-BHC	4.672	3.962	72327658	165.0E6	35.227	40.663
8)	B Heptachlo...	5.564	4.544	87843023	176.3E6	50.389	52.140
9)	A Endosulfan I	5.951	4.910	75600800	146.0E6	46.907	46.262
10)	B gamma-Chl...	5.820	4.792	91010337	196.7E6	51.979	57.706
11)	B alpha-Chl...	5.897	4.855	95502580	176.0E6	55.289	53.145
12)	B 4,4'-DDE	6.076	5.050	74224915	177.0E6	46.628	54.405
13)	MA Dieldrin	6.228	5.172	81677519	185.0E6	50.605	55.793
14)	MA Endrin	6.460	5.444	69192671	161.1E6	50.031	56.502
15)	B Endosulfa...	6.688	5.741	69319364	144.7E6	50.182	54.903
16)	A 4,4'-DDD	6.601	5.602	56262538	127.5E6	48.052	49.223
17)	MA 4,4'-DDT	6.912	5.850	64216051	143.4E6	50.678	54.422
18)	B Endrin al...	6.819	5.920	53326498	109.8E6	50.428	53.232
19)	B Endosulfa...	7.051	6.143	60042637	131.4E6	46.886	48.860
20)	A Methoxychlor	7.392	6.427	30580255	80352644	46.302	58.676m#
21)	B Endrin ke...	7.540	6.642	62301068	142.0E6	45.087	46.878
22)	Mirex	7.999	6.817	52613446	105.5E6	50.109m	49.082m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092022\
Data File : PL077869.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Sep 2022 16:33
Operator : AR\AJ
Sample : N4703-02MS
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
ETGI-363MS

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 09/21/2022
Supervised By :Ankita Jodhani 09/21/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 21 02:53:23 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL091222.M
Quant Title : GC Extractables
QLast Update : Tue Sep 13 13:26:11 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

