

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092120\
 Data File : PL061636.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Sep 2020 15:34
 Operator : DD\AJ
 Sample : 50 PPB SPIKE
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 50 PPB SPIKE

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 21 18:22:42 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL090820.M
 Quant Title : GC Extractables
 QLast Update : Wed Sep 09 01:27:38 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
Target Compounds							
2) A	alpha-BHC	3.905	4.399	135.9E6	239.8E6	53.107	51.930
3) MA	gamma-BHC...	4.188	4.685	95666237	220.5E6	41.915	51.389
4) MA	Heptachlor	4.491	5.196	109.6E6	198.5E6	51.071	50.769
5) MB	Aldrin	4.739	5.503	100.8E6	185.2E6	52.521	49.010
6) B	beta-BHC	4.440	4.857	48714038	87176330	50.712	49.920
7) B	delta-BHC	4.644	5.074	116.6E6	183.5E6	53.814	57.536
8) B	Heptachlo...	5.181	5.891	94919704	158.3E6	50.943	53.029
9) A	Endosulfan I	5.520	6.249	87071296	151.1E6	50.610	47.093
10) B	gamma-Chl...	5.409	6.126	94712364	166.0E6	51.125	48.145
11) B	alpha-Chl...	5.467	6.199	94197744	163.7E6	50.849	47.034
12) B	4,4'-DDE	5.632	6.357	86625420	145.8E6	50.943	45.480
13) MA	Dieldrin	5.763	6.507	92870360	149.4E6	51.739	45.564
14) MA	Endrin	6.019	6.725	78840762	116.7E6	46.914	41.752
15) B	Endosulfa...	6.293	6.935	84187121	149.3E6	49.208	50.553
16) A	4,4'-DDD	6.147	6.849	78429799	126.8E6	54.270	49.393
17) MA	4,4'-DDT	6.388	7.148	69597732	100.8E6	50.687	40.290
18) B	Endrin al...	6.461	7.060	73523694	97133962	49.945	39.703
19) B	Endosulfa...	6.673	7.283	81255815	117.3E6	49.372	42.886
20) A	Methoxychlor	6.934	7.610	38294354	68389579	44.733	47.544
21) B	Endrin ke...	7.162	7.756	95420605	108.3E6	52.180	39.874
22)	Mirex	7.351	8.211	66420057	79056269	44.420	41.057

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092120\
Data File : PL061636.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Sep 2020 15:34
Operator : DD\AJ
Sample : 50 PPB SPIKE
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
50 PPB SPIKE

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 21 18:22:42 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL090820.M
Quant Title : GC Extractables
QLast Update : Wed Sep 09 01:27:38 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

