

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092125\
 Data File : PL097294.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 Sep 2025 11:15
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 20 01:18:04 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092025.M
 Quant Title : GC Extractables
 QLast Update : Fri Sep 19 05:40:25 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.526	2.823	223.2E6	321.8E6	49.301	48.937
28)	SA Decachlor...	9.000	7.983	157.1E6	272.8E6	48.011	47.923
Target Compounds							
2)	A alpha-BHC	3.974	3.328	350.6E6	498.3E6	50.177	50.047
3)	MA gamma-BHC...	4.301	3.660	329.9E6	464.7E6	49.892	49.892
4)	MA Heptachlor	4.892	4.007	329.3E6	457.4E6	49.365	49.157
5)	MB Aldrin	5.231	4.290	314.2E6	430.0E6	49.609	49.544
6)	B beta-BHC	4.489	3.956	134.5E6	197.3E6	49.341	48.566
7)	B delta-BHC	0.000	4.188	0	459.4E6	N.D.	49.977 #
8)	B Heptachlo...	5.652	4.791	288.2E6	389.4E6	50.149	49.091
9)	A Endosulfan I	6.033	5.162	255.5E6	368.0E6	49.310	47.888
10)	B gamma-Chl...	5.905	5.043	281.3E6	425.3E6	49.758	48.698
11)	B alpha-Chl...	5.985	5.107	278.5E6	409.0E6	49.037	49.273
12)	B 4,4'-DDE	6.156	5.296	236.7E6	380.4E6	49.683	49.495
13)	MA Dieldrin	6.305	5.427	271.9E6	403.6E6	49.595	49.355
14)	MA Endrin	6.532	5.701	227.9E6	362.3E6	49.403	48.233
15)	B Endosulfa...	6.745	5.993	235.2E6	343.9E6	47.823	48.373
16)	A 4,4'-DDD	6.665	5.848	188.0E6	318.5E6	49.862	49.390
17)	MA 4,4'-DDT	6.979	6.101	209.5E6	345.5E6	49.570	49.928
18)	B Endrin al...	6.874	6.171	152.2E6	249.4E6	49.361	46.727
19)	B Endosulfa...	7.107	6.395	201.2E6	333.5E6	49.291	48.681
20)	A Methoxychlor	7.452	6.673	102.9E6	182.8E6	49.334	48.482
21)	B Endrin ke...	7.586	6.899	212.3E6	369.0E6	50.015	49.002
22)	Mirex	8.065	7.089	169.0E6	288.3E6	47.599	48.224
24)	Chlordane-2	5.231f	0.000	314.2E6	0	1415.546	N.D. #
25)	Chlordane-3	5.905	5.043	281.3E6	425.3E6	344.891	475.632 #
26)	Chlordane-4	5.985	5.107	278.5E6	409.0E6	278.463	537.050 #
27)	Chlordane-5	0.000	5.993	0	343.9E6	N.D.	1153.446 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092125\
Data File : PL097294.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 Sep 2025 11:15
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 20 01:18:04 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092025.M
Quant Title : GC Extractables
QLast Update : Fri Sep 19 05:40:25 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.526 min
Delta R.T.: 0.000 min
Response: 223193353
Conc: 49.30 ng/ml

Instrument :
ECD_L
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 2.823 min
Delta R.T.: 0.000 min
Response: 321788539
Conc: 48.94 ng/ml




#2 alpha-BHC

R.T.: 3.974 min
Delta R.T.: 0.000 min
Response: 350609105
Conc: 50.18 ng/ml



#2 alpha-BHC


R.T.: 3.328 min
Delta R.T.: 0.000 min
Response: 498344903
Conc: 50.05 ng/ml



#3 gamma-BHC (Lindane)


R.T.: 4.301 min
Delta R.T.: -0.001 min
Response: 329940214
Conc: 49.89 ng/ml

Instrument :
ECD_L
ClientSampleId :




#3 gamma-BHC (Lindane)

R.T.: 3.660 min
Delta R.T.: 0.000 min
Response: 464724279
Conc: 49.89 ng/ml




#4 Heptachlor

R.T.: 4.892 min
Delta R.T.: -0.001 min
Response: 329267088
Conc: 49.36 ng/ml



#4 Heptachlor


R.T.: 4.007 min
Delta R.T.: 0.000 min
Response: 457414041
Conc: 49.16 ng/ml



#5 Aldrin


R.T.: 5.231 min
Delta R.T.: 0.000 min
Response: 314213972
Conc: 49.61 ng/ml

Instrument :
ECD_L
ClientSampleId :




#5 Aldrin

R.T.: 4.290 min
Delta R.T.: 0.000 min
Response: 430004418
Conc: 49.54 ng/ml




#6 beta-BHC

R.T.: 4.489 min
Delta R.T.: 0.000 min
Response: 134513033
Conc: 49.34 ng/ml



#6 beta-BHC


R.T.: 3.956 min
Delta R.T.: 0.000 min
Response: 197276081
Conc: 48.57 ng/ml



#7 delta-BHC


R.T.: 0.000 min
Exp R.T.: 4.735 min
Response: 0
Conc: N.D.

Instrument :
ECD_L
ClientSampleId :




#7 delta-BHC

R.T.: 4.188 min
Delta R.T.: 0.000 min
Response: 459401721
Conc: 49.98 ng/ml




#8 Heptachlor epoxide

R.T.: 5.652 min
Delta R.T.: 0.000 min
Response: 288202649
Conc: 50.15 ng/ml



#8 Heptachlor epoxide


R.T.: 4.791 min
Delta R.T.: -0.001 min
Response: 389364629
Conc: 49.09 ng/ml



#9 Endosulfan I


R.T.: 6.033 min
Delta R.T.: 0.000 min
Response: 255493192
Conc: 49.31 ng/ml

Instrument :
ECD_L
ClientSampleId :




#9 Endosulfan I

R.T.: 5.162 min
Delta R.T.: 0.000 min
Response: 367952619
Conc: 47.89 ng/ml




#10 gamma-Chlordane

R.T.: 5.905 min
Delta R.T.: 0.000 min
Response: 281306901
Conc: 49.76 ng/ml



#10 gamma-Chlordane


R.T.: 5.043 min
Delta R.T.: -0.001 min
Response: 425318809
Conc: 48.70 ng/ml



#11 alpha-Chlordane


R.T.: 5.985 min
Delta R.T.: -0.001 min
Response: 278525424
Conc: 49.04 ng/ml

Instrument :
ECD_L
ClientSampleId :




#11 alpha-Chlordane

R.T.: 5.107 min
Delta R.T.: 0.000 min
Response: 408987881
Conc: 49.27 ng/ml




#12 4,4'-DDE

R.T.: 6.156 min
Delta R.T.: 0.000 min
Response: 236727202
Conc: 49.68 ng/ml



#12 4,4'-DDE


R.T.: 5.296 min
Delta R.T.: -0.001 min
Response: 380355999
Conc: 49.50 ng/ml



#13 Dieldrin


R.T.: 6.305 min
Delta R.T.: 0.000 min
Response: 271886176
Conc: 49.59 ng/ml

Instrument :
ECD_L
ClientSampleId :




#13 Dieldrin

R.T.: 5.427 min
Delta R.T.: -0.001 min
Response: 403567044
Conc: 49.36 ng/ml




#14 Endrin

R.T.: 6.532 min
Delta R.T.: 0.000 min
Response: 227904770
Conc: 49.40 ng/ml



#14 Endrin


R.T.: 5.701 min
Delta R.T.: -0.002 min
Response: 362254023
Conc: 48.23 ng/ml



#15 Endosulfan II


R.T.: 6.745 min
Delta R.T.: 0.000 min
Response: 235211105
Conc: 47.82 ng/ml

Instrument :
ECD_L
ClientSampleId :




#15 Endosulfan II

R.T.: 5.993 min
Delta R.T.: -0.002 min
Response: 343891588
Conc: 48.37 ng/ml




#16 4,4'-DDD

R.T.: 6.665 min
Delta R.T.: 0.000 min
Response: 188008206
Conc: 49.86 ng/ml



#16 4,4'-DDD


R.T.: 5.848 min
Delta R.T.: -0.001 min
Response: 318519128
Conc: 49.39 ng/ml



#17 4,4'-DDT


R.T.: 6.979 min
Delta R.T.: 0.000 min
Response: 209540798
Conc: 49.57 ng/ml

Instrument :
ECD_L
ClientSampleId :




#17 4,4'-DDT

R.T.: 6.101 min
Delta R.T.: 0.000 min
Response: 345524237
Conc: 49.93 ng/ml




#18 Endrin aldehyde

R.T.: 6.874 min
Delta R.T.: 0.000 min
Response: 152171808
Conc: 49.36 ng/ml



#18 Endrin aldehyde


R.T.: 6.171 min
Delta R.T.: -0.001 min
Response: 249403586
Conc: 46.73 ng/ml



#19 Endosulfan Sulfate


R.T.: 7.107 min
Delta R.T.: 0.000 min
Response: 201240347
Conc: 49.29 ng/ml

Instrument :
ECD_L
ClientSampleId :




#19 Endosulfan Sulfate

R.T.: 6.395 min
Delta R.T.: 0.000 min
Response: 333457717
Conc: 48.68 ng/ml




#20 Methoxychlor

R.T.: 7.452 min
Delta R.T.: 0.000 min
Response: 102880788
Conc: 49.33 ng/ml



#20 Methoxychlor


R.T.: 6.673 min
Delta R.T.: 0.000 min
Response: 182751500
Conc: 48.48 ng/ml



#21 Endrin ketone


R.T.: 7.586 min
Delta R.T.: 0.000 min
Response: 212281708
Conc: 50.02 ng/ml

Instrument :
ECD_L
ClientSampleId :




#21 Endrin ketone

R.T.: 6.899 min
Delta R.T.: 0.000 min
Response: 368999132
Conc: 49.00 ng/ml




#22 Mirex

R.T.: 8.065 min
Delta R.T.: 0.000 min
Response: 168992105
Conc: 47.60 ng/ml



#22 Mirex


R.T.: 7.089 min
Delta R.T.: 0.000 min
Response: 288318237
Conc: 48.22 ng/ml



#24 Chlordane-2


R.T.: 5.231 min
Delta R.T.: 0.028 min
Response: 314213972
Conc: 1415.55 ng/m

Instrument :
ECD_L
ClientSampleId :




#24 Chlordane-2

R.T.: 0.000 min
Exp R.T. : 4.412 min
Response: 0
Conc: N.D.




#25 Chlordane-3

R.T.: 5.905 min
Delta R.T.: -0.002 min
Response: 281306901
Conc: 344.89 ng/ml



#25 Chlordane-3


R.T.: 5.043 min
Delta R.T.: -0.002 min
Response: 425318809
Conc: 475.63 ng/ml



#26 Chlordane-4


R.T.: 5.985 min
Delta R.T.: -0.006 min
Response: 278525424
Conc: 278.46 ng/ml

Instrument :
ECD_L
ClientSampleId :




#26 Chlordane-4

R.T.: 5.107 min
Delta R.T.: 0.000 min
Response: 408987881
Conc: 537.05 ng/ml




#27 Chlordane-5

R.T.: 0.000 min
Exp R.T. : 6.828 min
Response: 0
Conc: N.D.



#27 Chlordane-5


R.T.: 5.993 min
Delta R.T.: -0.010 min
Response: 343891588
Conc: 1153.45 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.000 min
Delta R.T.: 0.000 min
Response: 157103382
Conc: 48.01 ng/ml

Instrument :
ECD_L
ClientSampleId :



#28 Decachlorobiphenyl

R.T.: 7.983 min
Delta R.T.: 0.000 min
Response: 272755219
Conc: 47.92 ng/ml