

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\
 Method File : LTX092218.M
 Title : GC Extractables
 Last Update : Sat Sep 22 02:44:37 2018
 Response Via : Initial Calibration

Calibration Files

500 =PL040208.D 1000 =PL040206.D 250 =PL040209.D
 750 =PL040207.D 50 =PL040210.D

	Compound	500	1000	250	750	50	Avg		%RSD

1) SA	Tetrachloro-m-xylene	7.513	7.593	7.445	7.274	7.704	7.506	E6	2.15
2)	Toxaphene-1	5.533	6.188	5.395	5.531	5.455	5.620	E4	5.73
3)	Toxaphene-2	1.970	1.994	1.990	2.182	2.198	2.067	E5	5.47
4)	Toxaphene-3	1.412	1.541	1.406	1.444	1.323	1.425	E5	5.50
5)	Toxaphene-4	6.815	7.730	6.796	7.225	7.208	7.155	E4	5.33
6)	Toxaphene-5	2.455	2.484	2.468	2.595	2.529	2.506	E5	2.27
7) SA	Decachlorobiphenyl	8.576	8.577	8.795	8.352	9.690	8.798	E6	5.94

Signal #2 Calibration Files

500 =PL040208.D 1000 =PL040206.D 250 =PL040209.D
 750 =PL040207.D 50 =PL040210.D

	Compound	500	1000	250	750	50	Avg		%RSD

1) SA	Tetrachloro-m-xylene	3.225	3.294	3.213	3.190	3.250	3.234	E6	1.23
2)	Toxaphene-1	1.020	0.980	1.054	1.060	1.316	1.086	E4	12.20
3)	Toxaphene-2	3.032	3.149	3.092	3.055	3.512	3.168	E4	6.23
4)	Toxaphene-3	9.507	9.864	9.591	8.559	9.173	9.339	E4	5.36
5)	Toxaphene-4	5.660	5.865	5.701	5.793	5.685	5.741	E4	1.50
6)	Toxaphene-5	2.222	2.325	2.253	2.210	2.322	2.266	E4	2.41
7) SA	Decachlorobiphenyl	3.119	3.010	3.278	2.997	3.722	3.225	E6	9.29

(#) = Out of Range