

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092220\
 Data File : PL061652.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Sep 2020 14:37
 Operator : DD\AJ
 Sample : L4100-06
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 021

Manual Integrations
 APPROVED

Ankita
 9/23/2020 11:27:46 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 23 03:38:41 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL090820.M
 Quant Title : GC Extractables
 QLast Update : Wed Sep 09 01:27:38 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.458	4.007	38183907	67956194	22.140	22.093
28)	SA Decachlor...	8.210	9.094	20336250	28592452	12.190	12.092
Target Compounds							
12)	B 4,4'-DDE	5.619	6.351	24382455	41713329	14.339	13.013
16)	A 4,4'-DDD	6.135	6.840	1388066	1322359	0.960m	0.515m#
17)	MA 4,4'-DDT	6.375	7.143	11040866	19942641	8.041	7.967

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092220\
Data File : PL061652.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Sep 2020 14:37
Operator : DD\AJ
Sample : L4100-06
Misc :
ALS Vial : 16 Sample Multiplier: 1



Instrument :
ECD_L
ClientSampled :
021


Manual Integrations
APPROVED

Ankita
9/23/2020 11:27:46 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 23 03:38:41 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL090820.M
Quant Title : GC Extractables
QLast Update : Wed Sep 09 01:27:38 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene

R.T.: 3.458 min
Delta R.T.: -0.001 min
Response: 38183907
Conc: 22.14 ng/ml

Instrument :
ECD_L
ClientSampleId :
021


Manual Integrations
APPROVED

Ankita
9/23/2020 11:27:46 AM




#1 Tetrachloro-m-xylene

R.T.: 4.007 min
Delta R.T.: -0.001 min
Response: 67956194
Conc: 22.09 ng/ml




#12 4,4'-DDE

R.T.: 5.619 min
Delta R.T.: -0.002 min
Response: 24382455
Conc: 14.34 ng/ml



#12 4,4'-DDE

R.T.: 6.351 min
Delta R.T.: 0.000 min
Response: 41713329
Conc: 13.01 ng/ml




#16 4,4'-DDD

R.T.: 6.135 min
Delta R.T.: 0.000 min
Response: 1388066
Conc: 0.96 ng/ml m

Instrument :
ECD_L
ClientSampleId :
021


Manual Integrations
APPROVED

Ankita
9/23/2020 11:27:46 AM




#16 4,4'-DDD

R.T.: 6.840 min
Delta R.T.: -0.003 min
Response: 1322359
Conc: 0.51 ng/ml m




#17 4,4'-DDT

R.T.: 6.375 min
Delta R.T.: -0.001 min
Response: 11040866
Conc: 8.04 ng/ml



#17 4,4'-DDT

R.T.: 7.143 min
Delta R.T.: 0.000 min
Response: 19942641
Conc: 7.97 ng/ml




#28 Decachlorobiphenyl

R.T.: 8.210 min
Delta R.T.: 0.000 min
Response: 20336250
Conc: 12.19 ng/ml

Instrument :
ECD_L
ClientSampleId :
021

Manual Integrations
APPROVED

Ankita
9/23/2020 11:27:46 AM



#28 Decachlorobiphenyl

R.T.: 9.094 min
Delta R.T.: 0.000 min
Response: 28592452
Conc: 12.09 ng/ml