

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092320\
 Data File : PL061695.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Sep 2020 14:51
 Operator : DD\AJ
 Sample : PSTDICV050
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 ICVPL092320

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 24 07:07:38 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092320.M
 Quant Title : GC Extractables
 QLast Update : Wed Sep 23 13:58:40 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.458	4.007	83994492	127.2E6	49.236	46.623
28)	SA Decachlor...	8.211	9.095	82526137	109.6E6	50.730	49.786
Target Compounds							
2)	A alpha-BHC	3.894	4.396	119.1E6	185.7E6	48.871	46.480
3)	MA gamma-BHC...	4.177	4.682	105.5E6	168.4E6	48.250	45.497
4)	MA Heptachlor	4.479	5.193	102.2E6	165.5E6	50.081	47.785
5)	MB Aldrin	4.726	5.499	91749099	155.7E6	50.850	48.060
6)	B beta-BHC	4.428	4.854	45966424	72276901	50.329	47.867
7)	B delta-BHC	4.632	5.071	101.8E6	130.8E6	49.639	51.345
8)	B Heptachlo...	5.169	5.886	86702844	125.0E6	49.662	50.422
9)	A Endosulfan I	5.507	6.244	80990487	129.9E6	49.981	47.022
10)	B gamma-Chl...	5.396	6.121	86599677	138.6E6	49.865	47.196
11)	B alpha-Chl...	5.455	6.194	85819502	138.7E6	49.405	46.179
12)	B 4,4'-DDE	5.620	6.352	79272513	130.4E6	50.439	47.765
13)	MA Dieldrin	5.751	6.502	83282135	133.3E6	49.517	46.976
14)	MA Endrin	6.007	6.720	76950579	113.0E6	48.360	46.321
15)	B Endosulfa...	6.280	6.929	73398294	111.1E6	45.397	42.988
16)	A 4,4'-DDD	6.135	6.843	67586126	105.8E6	51.046	47.578
17)	MA 4,4'-DDT	6.376	7.143	66479117	109.3E6	52.699	49.908
18)	B Endrin al...	6.448	7.055	60608348	91936898	43.101	42.952
19)	B Endosulfa...	6.661	7.278	75823197	112.0E6	49.012	47.828
20)	A Methoxychlor	6.922	7.603	41201874	60925298	50.960	48.354
21)	B Endrin ke...	7.151	7.750	86880948	114.7E6	50.837	48.872
22)	Mirex	7.340	8.206	70217182	88742436	50.087	50.366

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092320\
Data File : PL061695.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Sep 2020 14:51
Operator : DD\AJ
Sample : PSTDICV050
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
ICVPL092320

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 24 07:07:38 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092320.M
Quant Title : GC Extractables
QLast Update : Wed Sep 23 13:58:40 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

