

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092418\
 Data File : PL040249.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 24 Sep 2018 15:53
 Operator : AJ\SJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Manual Integrations
 APPROVED

Sohil
 9/25/2018 3:49:39 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 25 00:52:03 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092218.M
 Quant Title : GC Extractables
 QLast Update : Sat Sep 22 05:31:14 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.249	3.842	347.9E6	159.9E6	45.730	48.959
28)	SA Decachlor...	7.873	8.793	384.0E6	145.5E6	43.360	45.487
Target Compounds							
2)	A alpha-BHC	3.673	4.216	548.9E6	240.3E6	45.690	50.394
3)	MA gamma-BHC...	3.945	4.493	529.3E6	222.1E6	47.957	50.292
4)	MA Heptachlor	4.223	4.994	529.9E6	223.8E6	46.118	50.298
5)	MB Aldrin	4.457	5.292	485.8E6	204.8E6	45.963	50.183m
6)	B beta-BHC	4.200	4.656	196.7E6	95616091	44.759	48.436
7)	B delta-BHC	4.389	4.866	513.0E6	206.4E6	45.971	45.634
8)	B Heptachlo...	4.890	5.670	450.6E6	187.6E6	44.323	50.801
9)	A Endosulfan I	5.215	6.019	435.1E6	177.4E6	43.944m	50.066
10)	B gamma-Chl...	5.110	5.900	478.9E6	194.7E6	44.344	50.568m
11)	B alpha-Chl...	5.165	5.974	465.2E6	191.3E6	44.631	50.297
12)	B 4,4'-DDE	5.334	6.132	469.0E6	190.9E6	44.504	50.375
13)	MA Dieldrin	5.453	6.272	475.3E6	187.4E6	44.589	51.459
14)	MA Endrin	5.702	6.484	429.2E6	162.2E6	43.595m	50.804
15)	B Endosulfa...	5.978	6.688	424.5E6	158.8E6	40.865	51.041
16)	A 4,4'-DDD	5.842	6.611	386.5E6	152.1E6	44.729	47.504
17)	MA 4,4'-DDT	6.074	6.908	389.9E6	150.9E6	46.126	52.844
18)	B Endrin al...	6.145	6.810	336.8E6	136.0E6	43.503	49.614
19)	B Endosulfa...	6.355	7.032	393.8E6	149.2E6	43.983	50.368
20)	A Methoxychlor	6.620	7.364	202.1E6	69282869	43.118	45.484
21)	B Endrin ke...	6.832	7.492	437.7E6	157.0E6	44.787	53.567

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL092418\
Data File : PL040249.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Sep 2018 15:53
Operator : AJ\SJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Manual Integrations
APPROVED

Sohil
9/25/2018 3:49:39 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 25 00:52:03 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL092218.M
Quant Title : GC Extractables
QLast Update : Sat Sep 22 05:31:14 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

